

CODAN FRANCE Sarl.
49 Rue de Rohrwiler
BP40073
67242 Bischwiller CEDEX
Französische Republik

* TEST CERTIFICATE *

Product: STERITEX® V-Airvent Air Venting Needle with check valve, sterile

Quantity :	1.200	CODAN REF :	64.6008
LOT :	P84508-1	Invoice No :	201095
Steril.Date :	2020-08	Expiry Date :	2025-06-28

Results

EN ISO 11135

Sterilization of health-care products

Passed

European Pharmacopoeia

The analysis of bacterial endotoxin was performed with the rFC test method to EP chapter 5.1.10

Passed

Date: 22 September 2020

This certificate has been produced electronically and is valid without signature.