

Certificate of Quality



Part Number	MFLX06509-26
Lot Number	28585281
Part Description	TBNG MFLEX TYGON E-LAB NO. 26 50FT
Expiration Date	08SEP2024
Country of Origin	USA

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- Compound:E-3603
- This product meets 21 CFR 176.170(c) under condition of use: E through H as described in Table 1 and 2.
- Product has been manufactured in accordance with Good Manufacturing Practices.
- A lot control system is used to assure traceabilities to all raw materials.
- The proprietary compound has been tested: USP Class VI Toxicity Tests and have passed accordingly.
- "EU Food # Complies with 1935/2004/EC and 10/2011/EU for food contact with many foods and beverages. "
- For complete compliance information and appropriate use instructions, please refer to the detailed document of compliance.
- Tubing specifications and/or characteristics may change past this date. It is solely up to the user to determine suitability.

Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.