

STERILE STAINLESS SURGICAL BLADE SPECIFICATIONS

Supplied in boxes of 100

Blades in this range:

6, 9, 10, 10A, 10R, 11, 11P, E/11, 12, 12D, 13, 14, 15, 15A, 15C, 15T, 16
Fit Handles No.3 Graduated, 3L, B3, No.7, No.9

18, 19, 20, 21, 22, 22A, 23, 24, 25, 25A, 26, 27
Fit Handles No.4 Graduated, 4L

COMPONENT	MATERIAL
BLADE	Stainless Steel <u>ANALYSIS:</u> CARBON: 0.60% - 0.70% SILICON: 0.50% max MANGANESE: 1% max CHROMIUM: 12% - 13.5% SULPHUR: 0.025% max PHOSPHORUS: 0.03% max NICKEL: 0.5% max
TOP FOIL	Printed Aluminum with Polyethylene
BOTTOM FOIL	Printed Aluminum blade shape with CPP or HDPE laminate
RUST INHIBITOR	Waxed VCI impregnated kraft paper
UNIT CONTAINER BOX	White lined chipboard with litho machine varnish

Products are manufactured in accordance with BS 7153-1:2001 and BS EN 27740 (where applicable).

All Components are manufactured in Europe.

All packets and boxes are marked with the current lot number and expiry dates for traceability; boxes are bar coded where applicable.

Products are sterilised by Cobalt 60 Gamma Irradiation by:-Swann-Morton (Services) Limited, Owlerton Green, Sheffield S6 2BJ