

Certificate of Conformity

VWR Nitrile Extra Light Glove

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 112-4193

Lot No: C25090321

Manufacture Date: Sep 2025

Expiry Date: Sep 2030

Size: XS

Sampling plan: In accordance to ISO 2859-1: 1999

Parameter	Specification	Result	Sampling size
1.Dimension			
Palm width	≤ 80	Min. 76 Max: 78	Pass
Length	Median : 240mm	Median : 244 mm	Pass
Cuff Thickness	Min. 0.05 mm	Min. 0.05 mm	Pass
Palm Thickness	Min. 0.06 mm	Min. 0.06 mm	Pass
Finger Thickness	Min. 0.07 mm	Min. 0.08mm	Pass
2. Functionality and Appearance test			
Barrier	G1 AQL 1.5	Pass	
Visual	G1 AQL 2.5	Pass	

Certificate of Conformity

Parameter	Specification	Result	Sampling size
3. Lab testing			
Tensile Strength (Un-aged) (Mpa)	N/A	N/A	N/A
Tensile Strength (Aged) (Mpa)	N/A	N/A	N/A
Elongation at Break (Un-aged) (%)	N/A	N/A	N/A
Elongation at Break (Aged) (%)	N/A	N/A	N/A
Force at Break (After challenge test)(N)	Median : 6	6.8	Pass
Powder (mg)	≤ 2 mg	0.14 mg	Pass

Disclaimer: VWR's Quality Management System (QMS) is ISO 9001:2015 certified as a wholesale distributor of finished products, ensuring effective oversight of finished product handling, traceability, and distribution processes. VWR is not a manufacturer and does not engage in manufacturing, co-manufacturing or processing activities such as design, packaging, labeling, testing or quality release. Consequently, the supplier or manufacturer's QMS certification may differ from VWR's.

VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. This product is intended for research use only. It is not for use in diagnostic procedures, therapeutic applications, or any human or clinical use.

This document has been produced electronically and is valid without a signature.

Date of issue: **October 2025**