Water

ChromAR HPLC

Material No: 6795-25
Effective from: 2013/03/14

Product Specification

Test	Specification
Residue after Evaporation	<= 0.0005 %
Specific Conductance (25°C)	<= 2 uS/cm
Gradient Elution Test	
254 nm	<= 0.002 AU
Ultraviolet Absorbance (1.00-cm cell vs. water):	
210 nm	<= 0.004
254 nm	<= 0.002

For Laboratory, Research or Manufacturing use.
Suitable for Liquid Chromatography and UV-Spectrophotometry

Barbara Rykowska
Quality Assurance Coordinator

For questions on this Product Specification please contact Technical Service at technical.service@avantorsciences.com
Avantor Performance Materials Poland S.A. Sowinskiego 11 Gliwice 44-101 Poland Tel: +48 322392000