

Certificate Of Processing

Systagenix
Wound Management Ltd.
Gargrave, BD23 3RX, UK
Tel: +44 (0) 1756 747200

The products listed below have been irradiated at Gargrave Sterilisation Services, Systagenix, Gargrave, UK.

Customer: VWR International Ltd

Irradiation Batch Number/Schedule: 30,911

Irradiation Lot Start Date/Time: 17-Dec-21 3:24:31AM

Irradiation Lot End Date/Time: 17-Dec-21 6:19:45AM

Number Of Interruptions: 0

Comment: Product Description: Ergo 1L Square

PO: 4514059273

Expiry: 28/11/2023

Inventory Detail

Finished Goods Code	Lot Number	Released Quantity	Delivered Dose (kGy)		Specified Dose Range (kGy)
			Minimum	Maximum	
VWRI331-0283	1000CG/25	20	22.6	31.0	15.0 - 40.0

Gargrave Sterilisation Services/Systagenix certifies that the materials listed above received the absorbed doses within the precision and accuracy of the dosimetry system employed.

Signature Manifest

Signed By Supervisor **Lee Barrett**

Signed On: 17-Dec-21 At: 8:22 AM GMT:0:00

Signed By QA **Sharon Allport**

Signed On: 17-Dec-21 At: 10:27 AM GMT:0:00

Document Content Revision: 1