



L3/278030

OKC Allergy Supplies, Inc.

a division of

ALK-Abello, Inc.

Certificate of Analysis

Product: **SEV - Sterile Empty Vial**
Lot number: **SEV2071221**
Vial size: **20mL-20mm**
Manufacture date: **7/12/2021**
Part Number: **SEV20**

Quality Release Testing		
Assay	Specifications	Results
Appearance - Product	Essentially free of visible particulate	Pass
Appearance - Packaging	No apparent leakage or physical alteration	Pass
Sterility, USP <71>	Sterile	Conforms
Bacterial Endotoxins, USP <85>	≤ 0.25 EU/mL	≤ 0.0050 EU/mL

Release Date	Expiration Date
7/29/2021	07 - 2025

Quality Assurance Approval	
By / Date:	 08-05-2021

Original Certificate

L31278030

**OKC Allergy Supplies, Inc.**

a division of

ALK-Abello, Inc.

Certificate of Manufacture

Product: **SEV - Sterile Empty Vial**
Lot number: **SEV2011221**
Vial size: **10mL-20mm**

Manufacture date: **1/12/2021**
Dosage Form: **Empty, Sterile**
Manufacturing Document: **VMR-017**

Manufacturing Firm and Address: **OKC Allergy Supplies, Inc.
1005 S.W. Second Street
Oklahoma City, OK 73109**

Packaging Firm and Address: **OKC Allergy Supplies, Inc.
1005 S.W. Second Street
Oklahoma City, OK 73109**

"I hereby certify that the above information is authentic and accurate. The batch of product has been formulated / manufactured, including packaging and quality control, at the above-mentioned recognized facility in full compliance with the cGMP requirements of the local regulatory authority. The batch processing, packaging, and analysis records were reviewed and found to be in compliance with cGMP and all specifications and protocols for manufacturing at OKC Allergy Supplies, Inc."

Release Date	Expiration Date
1/28/2021	01 - 2025

Quality Assurance Approval	
By / Date:	<i>Car</i> <i>01-29-21</i>

Original Certificate