

Certificate of Conformity

VWR Glove, Nitrile Light, 240 mm, L

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 112-2373

Lot No: 2020-11 L069796 2011

Size: L

Sampling plan: In accordance to ISO 2859-1: 1999

Parameter	Specification	Result	Sampling size
1.Dimension			
Palm width	106 ± 5	Min. 106 Max: 109	Pass
Length	Median : 240 mm	Median : 247 mm	Pass
Cuff Thickness	Min. 0.10 mm	Min. 0.10 mm	Pass
Palm Thickness	Min. 0.11 mm	Min. 0.12 mm	Pass
Finger Thickness	Min. 0.15 mm	Min. 0.16 mm	Pass
2. Functionality and Appearance test			
Barrier	G1 AQL 1.5	Pass	
Visual	G1 AQL 2.5	Pass	

Certificate of Conformity

Parameter	Specification	Result	Sampling size
3. Lab testing			
Tensile Strength (Un-aged) (Mpa)	Min : 18 Mpa	24.9	Pass
Tensile Strength (Aged) (Mpa)	Min : 18 Mpa	24.1	Pass
Elongation at Break (Un-aged) (%)	Min : 500 %	560	Pass
Elongation at Break (Aged) (%)	Min : 400 %	480	Pass
Force at Break (After challenge test)(N)	Median : 6 N	9.8	Pass
Powder (mg)	≤ 2 mg	0.30 mg	Pass

This document has been produced electronically and is valid without a signature

Date of issue: **June 2022**