

FREUDENBERG MEDICAL



September 17, 2021

Freudenberg Medical Europe
2-8 Liebigstrasse
67661 Kaiserslautern
Germany

PO# 4500025063

QTY. 6 EA

PRODUCT CERTIFICATION

FREUDENBERG P/N DESCRIPTION FREUDENBERG L/N

70-400-07

0000175987

**¾" SILICONE GASKET
DOW CORNING, SILASTIC Q7-4780 MATERIAL**

<u>MATERIAL</u>	<u>FM P/N</u>	<u>FM L/N</u>
Silicone	30-011-01	0000164668
Elastomer	30-012-01	0000164669

Freudenberg Medical certifies that our silicone gaskets are manufactured following Good Manufacturing Practices (GMP) in accordance with the new Quality System Regulation (QSR) in a FDA registered medical device facility. Freudenberg Medical has been certified to meet ISO 13485 quality system. Material manufacturer states that these materials meet or exceed USP Class VI biocompatibility requirements and FDA Regulations CFR 177.2600 for contact with food and drugs. These platinum-cured silicone products have full traceability to raw materials and contain no peroxide catalysts.

Sincerely,

A handwritten signature in cursive script that reads "Robert Lodevico".

Robert Lodevico / Shipping Clerk

Freudenberg Medical
Material Number Batch Number

45639007-0000175987

(7040007 REV. B)

FREUDENBERG MEDICAL, LLC
1110 Mark Avenue
Carpinteria, CA 93013

Phone: +1 805 684 3304

Fax: +1 805 684 1934

www.freudenbergmedical.com