Certificate of Conformity

VWR Rack for Centrifuge Tube

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

Cat. No.	Lot no.	For Tube	Hole Ø	No. of holes	W×D×H	Colour	Pk
525-0971	J338107	5 ml	13 - 16 mm	40 (4×10)	260×113×105 mm	White	1

Material: Polypropylene, 20% Talc Filled White
Colorant White Polypropylene, FDA Grade

This document has been produced electronically and is valid without a signature

Date of issue: **August 2020**

VWR International bv, Researchpark Haasrode 2020, Geldenaaksebaan 464, 3001 Leuven, Belgium
http://www.vwr.com