

Certificate of Conformity

VWR Cleanroom Lab Coat, VWR Basic, SMS, Disposable

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 113-1125

Lot No: A3202254

Material: SMS 50g/m² – Light Blue

Date of Manufacture: 08th Jan 2021

Quality System Compliance: Products are manufactured under the ISO 9001:2015 & ISO 13485:2016 standard. Products are Inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

QC Testing:

Inspected according to AQL level II 4.0 sampling plan per MIL-STD-105E
Break Tensile properties (N): MD=105.48 CD=112.24 (ASTM D5034)
Elongation rate (%): MD=90.02 CD=118.34 (ASTM D5034)

BSE/TSE Statement: The above product contains no ingredients of animal origin and no material derived from, or exposed to animals affected by or under quarantine for Transmitting, Animal Spongiform Encephalopathy (TSE)/Bovine Spongiform Encephalopathy(BSE)

Other Characteristics: Latex Free

This document has been produced electronically and is valid without a signature

Date of Issue: **Feb 2022**