

Buffer Solution
Baker Analyzed
pH 4.00 [Phthalate]



Material No: 5606.1000
Effective from: 2016/01/29

Product Specification

Test	Specification
pH at 25°C	3.98 – 4.02

For Laboratory, Research or Manufacturing use.

Temperature (°C)

0 4.00 pH
10 4.00 pH
20 4.00 pH
30 4.01 pH
40 4.03 pH
50 4.05 pH
60 4.08 pH
70 4.12 pH
80 4.16 pH
90 4.21 pH

This buffer is traceable to NIST Standard Reference Buffer.
Contains Potassium Hydrogen Phthalate.
Contains 0.01% Sodium Benzoate.

Barbara Rykowska
Quality Assurance Coordinator

For questions on this Product Specification please contact Technical Service at technical.service@avantorsciences.com
Avantor Performance Materials Poland S.A. Sowinskiego 11 Gliwice 44-101 Poland Tel: +48 322392000