



STERIS Corporation Sterility Assurance
Steraffirm BOWIE DICK TEST PACK (EQC003) CERTIFICATE OF PERFORMANCE
Exhibit CB

CATALOG NUMBER: EQC003

PRODUCT NAME: Steraffirm Bowie-Dick Test Pack

LOT NUMBER: 12335806

EXPIRATION DATE: 06/01/2028

DATE OF MANUFACTURE: 10/30/2025

MANUFACTURE LOCATION: 6515 Hopkins Road, Mentor, OH 44060

This document certifies that Steraffirm Bowie Dick Test Pack 121° C/8 minutes. (P/N EQC003) meets STERIS Corporation's Quality Assurance Specifications and conforms to: ANSI/AAMI/ ISO 11140-4, Type 2 Indicator.

Quality Systems Representative: _____

Date: 05 Nov 2025

Reviewed By: _____

Date: 05 Nov 25

LIMITATION OF LIABILITY AND INDEMNITY

Nothing in this Certificate of Performance shall, or is intended to, alter, expand, or diminish the terms and conditions of sale governing your purchase of the Biological Indicator Product from STERIS. In no event, whether as a result of breach of warranty, or tort (including negligence and strict liability) shall STERIS or its suppliers be liable as a result of any statement or information contained in this Certificate of Performance. In addition, STERIS shall not be liable for any consequential or incidental damages including, without limitation, loss of use or damage to your products or equipment, cost of substitute products, or down time costs, allegedly caused by the Biological Indicator Products. The responsibility of STERIS for damages due to injuries or death caused by the Biological Indicator Product shall be limited to that portion of such damages as might be attributable to the negligence or strict liability or other tortuous conduct of STERIS.