

YANGZHOU MEDLINE INDUSTRY CO., LTD.

No. 108, Jinshan Road, Economic Development Zone, Yangzhou, China

To whom it may concern,

CERTIFICATE OF CONFORMITY

The products listed as enclosed had been manufactured according with ISO 13485:2016 and China GMP requirements.

(pls see list as enclosed)

I. Visual Inspection: conform;

II. Functional Inspection: Conform;

III. Sterilization: Conform;

IV. LAL Test (Absence of bacterial endotoxins): Conform;

The visual and functional inspections were completed in accordance with Terumo specifications based on EN/ISO standards.

Sterilization method: EO Sterilization

Sterilization parameters are determined and revalidated periodically according to EN ISO 11135-1.

Sterility Assurance Level (SAL) is 10^{-6} .

The absence of bacterial endotoxins was tested in accordance with USP specifications.

Tanwei
Manager
Quality Department
YANGZHOU MEDLINE INDUSTRY CO., LTD.



YANGZHOU MEDLINE INDUSTRY CO., LTD.

No. 108, Jinshan Road, Economic Development Zone, Yangzhou, China

2101TE&2102TE (TCH21063)

Product code	Lot No	Sterilization Date	Exp. Date
MDSS03LE	20210405T	2021-04-05	2026-04-04
MDSS03SE	20210406T	2021-04-06	2026-04-05
MDSS05S2138E	20210410T	2021-04-10	2026-04-09
MDSS01S2516E	20210505T	2021-05-05	2026-05-04
MDSS01S2613E	20210506T	2021-05-06	2026-05-05

