Certificate of Conformity

Catalogue Number	40015-0011
Description	Universal VWR 0,5 ml steril
Lot no.	1040701/1
Country of Origin	Manufactured in Germany
Date of Issue (dd/mm/yyyy)	12/01/2023
Production date	2022/11
Expiry date	2027/11

Quality System Compliance

Products are manufactured in **ISO 13485:2016** and **ISO 9001:2015** certified plant. Products are inspected and controlled in accordance with the current applicable product specification.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

Material	Tips: Polyethylene
Manufacturing Environment	Robotic tips are manufactured under clean room conditions according to DIN EN ISO 14644-1:2016-06.
Sterilization	Products are EB (Electron beam) irradiated and dose released upon ISO 11137-2:2013 recommended practices in effect at the time of validation meet the requirement of 10^{-6} SAL.
ATP Assay	Tested with detection level < $1*10^{-12}$
PCR inhibitors Free	Product yield equivalent to control < 10 amplifiable targets
Human DNA Free	Tested with detection level < 1 pg
DNase & RNase Free	Tested with detection level < $1*10^{-6}$ (DNase) and < $1*10^{-9}$ (RNase)
BSE/TSE	No use of any raw materials produced from or substances derived from animal origin. The manufacturing process of the product does not intentionally use any ingredient of animal origin and no material derived from or exposed to animals affected by.
Cytotoxicity	Not applicable

Non-Pyrogenic Statement	Not applicable
Latex Statement	Not applicable
BPA Statement	Not applicable
DEHP Statement	Not applicable
Endotoxine Free	Tested with detection level < 0,02 EU
RoHS	The raw materials used are compliant with the currently valid directive 2011/65/EU and 2002/95/EG (RoHS).
REACH Statement	The product does not contain any "SVHC" ingredients, in accordance with the currently valid regulation (EC) No 1907/2006 on REACH.
Storage Conditions	Keep dry and away from sunlight
Shelf Life	5 years

Disclaimer: Ritter states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.