1,2,4-Trichlorobenzene
HPLC
Suitable for Use in Gel Permeation Chromatography

Material No.: 9444-05
Revision No.: 0

Product Specification

Test	Specification
Assay ($C_6H_3Cl_3$) (by GC)	≥ 99 %
Ultraviolet Absorbance (1.00-cm cell vs. water) - 350 nm	≤ 0.15
Ultraviolet Absorbance (1.00-cm cell vs. water) - 375 nm	≤ 0.05
Ultraviolet Absorbance (1.00-cm cell vs. water) - 400 nm	≤ 0.010
Ultraviolet Absorbance (1.00-cm cell vs. water) - UV Cut-off, nm	≤ 310
Residue after Evaporation	≤ 5 ppm
Water (H_2O)(by Karl Fischer titrn)	≤ 0.01 %

For Laboratory,Research,or Manufacturing Use

Packaging Site: Paris Mfg Ctr & DC

Jamie Ethier
Vice President Global Quality

For questions on this Product Specification please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone 610.386.1700