

Test Report – Eppendorf Biopur –

Combitips advanced[®] 2.5 mL



100 Pcs.
U.S. Pat. www.eppendorf.com/lp
Eppendorf BIOPUR

LOT K194692H

2026-02

STERILE R

Order no.: 0030 089.650
Eppendorf AG
22331 Hamburg · Germany
Email: eppendorf@eppendorf.com

Cat. no.: 0030089650
Distributed in North America by
Eppendorf North America, Inc. · USA
Email: info@eppendorf.com

- **Human DNA**
Method: *real-time* PCR acc. to SOP 99880/V2
Limit: < 2 pg; less than one human cell
- **Bacterial DNA**
Method: *real-time* PCR acc. to SOP 100567/V1
Limit: < 50 fg; less than 10 *E. coli* cells
- **DNase**
Method: DNA Digestion acc. to SOP 100570/V1
Limit: Not detectable
(LOD: 1.0×10^{-6} Kunitz units)
- **RNase**
Method: RNA Digestion acc. to SOP 100571/V1
Limit: Not detectable
(LOD: 1.0×10^{-9} Kunitz units)
- **PCR inhibition**
Method: *real-time* PCR acc. to SOP 100559/V1
Limit: Less than 10 targets amplifiable
- **Sterility**
Method: Sterilization with SAL 10^{-6} by
☒ Irradiation according DIN EN ISO 11137:2015* or ☐ Ethylene oxide*
*Please see attached label for the standard method of sterilization for this product
Test method: USP, Ph. Eur. 8.0 / 2.6.12
- **Endotoxin**
Method: Kinetic-chromogenic LAL test (Ph. Eur. 8.0 / 2.6.14)
Limit: < 0.001 EU/mL
- **ATP**
Method: Pre-sterilization test / Bioburden test (Ph. Eur. 8.0 / 2.6.12)
Limit: < 5.5×10^{-12} mg

Customer: Eppendorf AG, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 16.03.21

Testing of the above described lot showed conformity within the limits of detection.

23.03.21

Date of release

Epp 2103 8408

Registration number

Dr. Weigel

Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.