



CERTIFICATE OF CONFORMANCE AND STERILIZATION

Date: 20-Feb-20

This certification is provided as full assurance that the following product code was manufactured in accordance with prescribed procedures and specifications.

Product/Material Number:	C2-1212IR
Product Description:	STERILESORB* C/P 12INx12IN
Lot Number:	174758
Irradiation Run Number:	2609006
Irradiation Dose Delivered:	33.1kGy to 42.4kGy
Manufacture Date:	02/2020
Expiration Date:	02/2022

The above product has been Gamma Irradiated to Sterility Assurance Level of 10^{-6} in accordance with the guidelines of the Association for the Advancement of Medical Instrumentation.

A handwritten signature in black ink, appearing to read "C. Inthavongsa".

Champagne Inthavongsa
Quality Supervisor
cinthavongsa@contecinc.com

Certificate of Processing

STERIGENICS 10811 Withers Cove Park Dr. Charlotte NC 28278
 TEL 704 588-6877 FAX 704 588-8355 www.sterigenics.com

R55480102

02/17/2020 17:24:32 GMT
 Page 1 of 1

Customer Name:	FG Clean Wipes- Chicopee, MA.	Processing Facility:	Charlotte Gamma	Work Order #	2609006
P.O.#	38508			Sales Order #	2432171
	25-50 kGy	LYM2, Gamma Treatment	Irradiation Date/Time:	02/14/2020 05:10:00 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	520	CA	C2-1212IR	Wiping Cloth	174181	NA
102.000	520	CA	C2-1212IR	Wiping Cloth	174758	NA
	1040	CA	Total			

Processing Summary

Op#	Operation Name	Location	Date/Time In (GMT)	Date/Time Out (GMT)
400.00	Receive Product	UNPROC	02/10/2020 16:04:00	02/11/2020 20:39:18
410.00	Prepare Product	UNPROC	02/11/2020 20:39:18	02/13/2020 02:27:55
420.00	Load Product - A123	UNPROC	02/13/2020 02:27:55	02/14/2020 05:10:00
430.00	Process Product	UNPROC	02/14/2020 05:10:00	02/14/2020 05:14:42
440.00	Unload Product	UNPROC	02/14/2020 05:14:42	02/14/2020 16:18:00
450.00	Collect Run Data	UNPROC	02/14/2020 16:18:00	02/14/2020 17:23:19

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	50.0 kGy	33.1 KGY	Pass	BPOSTON	02/14/2020 17:21:18 GMT
	Reason Code Test					BRANTT POSTON	
450.00	Maximum Dose	25.0 kGy	50.0 kGy	42.4 KGY	Pass	BPOSTON	02/14/2020 17:21:45 GMT
	Reason Code Test					BRANTT POSTON	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: WENDY KOBLISH
 Reason: Work Order Completions

Date: 02/17/2020 17:21:45 GMT