



MEDICAL PRODUCTS - CLEVELAND FACILITY
911 North Davis Avenue, Cleveland, MS 38732, Phone: 662.843.9421, Fax: 662.846.5987

Certificate of Analysis

Product: STERILE WATER FOR IRRIGATION, USP - 500 ML

Lot #: G125567

Code: 2F7113

Expiry: OCT 2019

Manufactured Date: 10/18/2016

Sterilization Date: 10/18/2016

Mixing Specification: 03-15-53-216

Issue Date: 02/26/2015

TEST NUMBER	TEST NAME	LIMIT	RESULT
11-20-15-001	Oxidizable Substances	Pass	Pass
11-21-16-003	pH at 25 deg. C	5.0 - 7.0	5.5
11-22-03-007	Conductivity Determination of Water (USP/EP)	NMT 5 uS/cm	1 uS/cm
20-03-07-006	Particle Analysis - Large Volume (Instrumental)		
	Particles >= 10 micron	NMT 25 particles/mL >=10 microns	1 ct/mL
	Particles >= 25 micron	NMT 3 particles/mL >=25 microns	0 ct/mL
13-01-10-313	Bacterial Endotoxin	< 0.25 EU/mL	<0.02 EU/mL
14-04-01-120	Sterility	Pass Parametric Release	Pass
PARAMETRIC RELEASE IN ACCORDANCE WITH U.S. FDA REQUIREMENTS			

THIS PRODUCT WAS MANUFACTURED ACCORDING TO CURRENT GMP,
FULLFILLS THE REQUIREMENTS OF THE MASTER PRODUCTION DOCUMENT AND TESTED TO USP COMPLIANCE

E. Soliz 11-1-16
Prepared By/Date

Ellen Soliz

Harter 11-1-16
Quality Management Approval /Date