

Catalog Number: 10143-618
Material description: Papanicolaou Stain OG-6
Lot Number: 7100
Use by/Expiration Date: 2019-04-10
Origin: West Deptford, NJ
Expiration Date: 2019-04-10
Storage: 15°C - 25°C

Certified Analysis

Test/Method	Unit of Measure	Specifications	Result
Absorbance: Visible Scan-(700- 400nm)	MA	270 – 350 @626 nm	327 MA@626nm

Physical Inspection:

Check for correct label c copy, lot number, expiration date, fill volume, absence of particulate matter and overall aesthetics

Pass X Fail _____

Release By: Samantha Moses

Quality Control, DS Date 06-27-2017

