



Certificate of Conformity

VWR Forcep, Sterile, Individually Wrapped

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 232-0191

Lot No: 20190320

Material: 100% High Clarity Polystyrene

Length: 12.5 cm

Quality System Compliance:

Products are manufactured under the EN ISO 13485:2016 standard. Products are inspected and controlled through the whole production process in accordance with current applicable product specifications and QC SOP. Inspection is performed according to AQL level S-2 1.5 sampling plan per ISO 2859-1:1999. Inspection records are reviewed and signed off by qualified personnel for product release.

Product Specifications:

This product is a Class I category device as defined by the FDA in 21CFR Parts 862-892.

Date of EO Sterilization: 03/20/2019

EO Sterilization Expiry Date (MM/DD/YYYY): 03/20/2022

EO Sterilization Process:

Product labeled as sterile is Ethylene Oxide (EO) irradiated according to AAMI/ ISO11135-1: 2007 recommended practices in effect at the time of validation. Expiration date (Shelf life) is based on product and packaging shelf life, provided the product wrapper is not damaged.



Certificate of Conformity

BSE/TSE Statement: The above product contains no, or has not been in contact with, any ingredients of animal origin and/or material derived from or exposed to animals affected by or under quarantine for Transmitting, Animal Spongiform Encephalopathy (TSE)/Bovine Spongiform Encephalopathy (BSE).

This document has been produced electronically and is valid without a signature

Date of Issue: **May 2019**