

Certificate of Conformity

VWR Catalogue Number	212-0426
Description	Test tube racks, with 24 (3x8) holes, 30 mm tube Diameter, 110 mm width, 285.5 mm Depth, 86.7 mm Height, Green
Lot Number	V46-090221
Date of Manufacture (yyyy-mm-dd)	2021-02-09
Country of Origin	Manufactured in India
Date of Issue (yyyy-mm-dd)	2021-12-20

Quality System Compliance

Products are manufactured under the **ISO 9001:2015** and **ISO 13485:2016** standards. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Test Name	Test Reference	Accuracy	Result
Autoclaving	QWI-QC-07	General Insp. Level-1, AQL -1.5	PASS
Dimensional	QWI-QC-23	General Insp. Level-1, AQL -1.5	Within Tolerance

Product Specifications

Material	30% Glass filled Nylon with color masterbatch.
Sterilization	Not Applicable.
ATP Assay	Not Applicable.
DNase & RNase Free	Not Applicable.
BSE/TSE	Intentional addition of raw materials made from substances of animal origin, but comply with European Regulation (EC) N° 1774/2002 and Directive 2000/6/EC. Also the requirements of paragraph 6.4 Tallow derivatives of EMEA/410/01 rev.3 1 July 2011 are met.

Cytotoxicity	Not Applicable.
Non-Pyrogenic Statement	Not Applicable.
Latex Statement	Not Applicable.
BPA Statement	Not Applicable.
DEHP Statement	Not Applicable.
RoHS	No substances (Lead, Cadmium, Mercury, Hexavalent Chromium (Cr6+), Poly Brominated Biphenyls (PBB), Poly Brominated Diphenyl ethers (PBDE), Bis(2-Ethyhexyl) Phthalate (DEHP), Benzyl Butyl Phthalate (BBP), Dibutyl Phthalate (DBP), Diisobutyl Phthalate (DIBP)) are used in manufacturing the raw materials and final product. No routinely analyse is performed.
REACH Statement	No substances as listed on >Substance of Very High Concern< candidate list published 2021-07-08 are used in manufacturing of the raw materials used to produce this product. No routinely analysis is performed.
Storage Conditions	Ambient temperature.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature.