

CERTIFICATE OF CONFORMITY & CERTIFICATE OF ANALYSIS

Brand	MICROFLEX®
Size(s)	L (8.5 - 9)
Material	Nitrile
Style	MidKnight® MK-296
Short Description	Distinctive, Black Nitrile Exam Glove
Lot Number	2408LC
Batch Number	24081414LC
Manufacturing Date	2024-08
Expiration Date	2027-08

CERTIFICATE OF CONFORMITY

Ansell certifies that the above product is manufactured in accordance with Ansell product specifications and is in conformity with the following standards or normative documents:

Standard/Normative Document	Result
FDA 21 CFR 177-2600	Conforms
Reach Regulation	Conforms

CERTIFICATE OF ANALYSIS

Property	Test Method	Specification	Result
Freedom From Holes (General Inspection Level I)			
Freedom from Holes - Final Release	EN455-1/ASTM D5151	G1, AQL 1.5	Conforms
Visual Inspection			
Visual Inspection-major	Internal test method	G-I, AQL 2.5	Conforms
Visual Inspection-minor	Internal test method	G1, AQL 4.0	Conforms
Physical Properties			
Material Standard	ASTM D6319	S2, AQL 4.0	Conforms
Dimension Test	ASTM D6319/ASTM D3767	S2, AQL 4.0	Conforms
Finger Thickness (mm)	ASTM D6319/ASTM D3767	0.12- 0.19	Conforms
Palm Thickness (mm)	ASTM D6319/ASTM D3767	0.09 - 0.13	Conforms
Cuff Thickness (mm)	ASTM D6319/ASTM D3767	0.09 -0.15	Conforms
Glove Length (mm)	ASTM D6319/ASTM D3767	Min. 240	Conforms
Palm Width (mm)	ASTM D6319/ASTM D3767	72-130	Conforms
Physical Properties	ASTM D412	S2, AQL 4.0	Conforms
Tensile Strength (MPa)	ASTM D412	Min. 16	Conforms
Ultimate Elongation (%)	ASTM D412	Min. 400	Conforms
Modulus at 300% elongation (MPa)	ASTM D412	Max. 9	Conforms
Additional Properties			

Property	Test Method	Specification	Result
Residual Powder Content (mg/glove)	ASTM D6124	Max. 2.0	Conforms
Protein Content (µg/g)	ASTM D5712	N/A	Conforms

Electronically Signed by:

Sandra Wessel-Jess

QA Manager - Americas

Reno, NV, United States

Date: 5 August 2025

For additional information visit us at www.ansell.com, or contact us at:

Europe, Middle East & Africa Region

Ansell Healthcare Europe N.V.

Riverside Business Park

Blvd International, 55

B-1070 Brussels,

Belgium

T: +32 (0) 2 528 74 00

F: +32 (0) 2 528 74 01

Australia & Asia Pacific Region

678 Victoria St,

Richmond VIC 3121

Australia

Tel: +61 1800 337 041

Fax: +611 800 803578

Email: anzcustomerservice@ansell.com

North America Region

Ansell Healthcare products LLC

111 S Wood Ave,

Suite 900 Iselin,

NJ 08830,

United States

T: +1 800 800 0444

F: +1 800 800 0445

CA Tel: +1 800 363 8340

Email: customerserviceus@ansell.com

Ansell Healthcare Products LLC

2301 Robb Drive,

Reno,

NV 89523,

United States

US Tel: +1 800 800 0444

US Fax: +1 800 800 0445

Email: customerserviceus@ansell.com

Latin America & Caribbean Region

Ansell Comercial México S.A. de C.V.

Blvd. Bernardo Quintana 7001-C,

Q7001 Torre 2, Suites 1304-1306 Centro Sur.

Querétaro, Qro. México C.P. 76090

Tel: +52 442 296 2050

Email: cslac@ansell.com

Ansell Brazil Ltda.

Rua das Figueiras, 474

4º andar - Bairro Jardim

Santo André - SP

CEP: 09080-300 – Brasil

Tel: +55 11 3356 3100

Email: sac@ansell.com