

CODAN



CODAN Medical ApS · Færgevej 4 · DK-4970 Rødby

CODAN FRANCE SARL
49 RUE DE ROHRWILLER,
BP 40073
67242 BISCHWILLER CEDEX
REPUBLIC OF FRANCE

CODAN Medical ApS

Telefon +45 54 65 74 11
www.codan.de
CVR-nr.: 16 27 20 78

TEST CERTIFICATE

Product : 1 ML CODAN TBC 0.01 GRAD. LUER. NEUTRAL GAS-PAPER
WITH FULL IN-LINE PRINT. INSERT IN BOX.

Quantity....: 51.200

CODAN REF: 62.1612

LOT: H89175-1

Invoice No.....: 407442

Steril.Date: 2020-07

DS/EN ISO 11138

Results

Biological Indicators for Ethylene Oxide Sterilization Passed

European Pharmacopoeia

2.6.14. Bacterial Endotoxins
(except oral tip one)

Passed

The validation of the sterilization process
was performed according to:

DS/EN ISO 11135-1

Sterilization of Health Care Products - Ethylene Oxide -
Requirements for Development, Validation and Routine Control
of a Sterilization Process for Medical Devices.

EN 556

Sterilization of Medical Devices -
Requirements for Medical Devices to be labelled "Sterile".

Date : Jul.09,20

This certificate has been issued
electronically and is valid
without signature or stamp.