

Thermo Fisher Scientific hereby certifies that the product identified below is manufactured according to the requirements of product and quality specifications as maintained in our quality management system which is compliant to ISO 13485 (BSI Certificate Number: FM 653694) or ISO 9001 (BSI Certificate Number: FM 743358) in Monterrey, NL, MEX.



Ruben Deschamps
Sr. Quality Manager
08/07/2026

The following information represents Product Certification for: Item#: **1203-0600**

Description: **BEAKER PMP GRADUATED; 600ML**

Lot#: **7474391010**

Manufactured: **06/05/2026**

Part Number	Description	Common Name	DMF#	Cytotoxicity	USP Class VI	FDA Compliance - 21 CFR
1-0211-90	BEAKER,PMP,600ML,W/O GRADS	COMPONENT PART				
8-0022-01	RESIN,PMP,INJ	POLYMETHYLPENTENE	3621	PASSED	PASSED	177.1520(a)(3)(ii)

If N/A appears in any of the columns above it means the information is not available. Any item listed as "COMPONENT PART" will show blank in the DMF#, Cytotoxicity, USP Class VI, and FDA Compliance Information columns.

If the word "PASSED" appears in the USP Class VI column next to the resin listing, this material has passed USP Class VI requirements, latest Volume, as part of our initial test approval protocol.

If the word "PASSED" appears in the Cytotoxicity column next to the resin listing, this material was tested and shown to be non-cytotoxic as part of our initial test approval protocol, using either mouse fibroblast L929 cells or the more sensitive human diploid lung cell lines WI-38 or MRC-5.

Beakers are designed to meet the ISO/DIS 7056 International Standards for Laboratory Plasticware Accuracy.