

Certificate of Quality



Part Number	MFLX06402-74
Lot Number	23329917
Part Description	SPT 50.125X.250 50
Expiration Date	28DEC2022
Country of Origin	USA

Inspection & Testing

- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Product Specifications & Compliance

- Compound: Silicone
- This material is not intended for invasive use.
- Meets or exceeds USP Class VI and USP 87 requirements, current version.
- Silicone products are compliant to the extractable limits specified in FDA 21 CFR 177.2600, Rubber Articles Intended for Repeated Use, paragraphs (d), (e), and (f).
- The materials used in the manufacture of this product that come in contact with the fluid pathway are either animal derivative content free or compliant as defined in specification EMEA 410.01 Rev.3.
- Shelf Life: 5 years from Date of Manufacture
- The component materials used in the production of the product above were manufactured in
a manner compliant with current good manufacturing practices.
- This product is manufactured in the USA.
- This product is not approved or intended for, and should not be used for medical, clinical, surgical or other patient oriented applications.
- It is the responsibility of the user to conduct tests that are deemed necessary to determine the suitability of this product for any particular use.

Storage

- Storage in original packaging, at ambient conditions and away from direct sunlight, is recommended.

Certificate Approval

- This certificate was issued and approved by the Avantor Fluid Handling Quality and Regulatory Departments and is valid without signature.