

Operating Instructions

Original Operating Instructions

Proline®

Mechanical Pipette

Single-channel Models | Multi-channel Models



1000134943



SARTORIUS

Contents

1	About These Instructions.....	5
1.1	Validity.....	5
1.2	Related Documents.....	6
1.3	Target Groups.....	6
1.4	Symbols Used.....	6
1.4.1	Warnings in Operation Descriptions.....	6
1.4.2	Other Symbols Used.....	6
2	Safety Instructions	7
2.1	Intended Use	7
2.1.1	Modifications to the Device	7
2.1.2	Repairs to the Device.....	7
2.2	Qualifications of Personnel	8
2.3	Functionality of the Device Parts.....	8
2.4	Risk of Injury During Prolonged Pipetting.....	8
3	Device Description.....	9
3.1	Device Overview.....	9
3.2	Pipette Tips	10
3.2.1	Safetyspace® Filter Tips	10
3.3	Adjustment Tool	11
3.4	Pipetting Practice.....	11
3.5	Initial Adjustment and Follow-up Adjustment	11
4	Installation	12
4.1	Scope of Delivery	12
4.2	Unpacking	12
4.3	Storing the Pipette When Not in Use	12
5	Commissioning.....	13
5.1	Checking the Replacement of the Tip Ejector	13
5.2	Adjusting the Volume (Only Pipettes with Variable Volumes).....	13
5.3	Attaching the Tip.....	14
5.4	Ejecting the Tip	14
5.5	Inserting or Removing a Safe-Cone Filter.....	15
5.5.1	Inserting a Safe-Cone Filter	15
5.5.2	Removing a Safe-Cone Filter.....	15
6	Operation	16
6.1	Considering Good Pipetting Practices.....	16
6.2	Forward Pipetting.....	17
6.3	Reverse Pipetting	17
6.3.1	Repetitive Reverse Pipetting	18
7	Cleaning and Maintenance.....	19
7.1	Testing Chemical Resistance.....	19
7.2	Cleaning the Outer Surface of the Pipette.....	19
7.3	Maintenance Schedule	20
7.4	Replacing the Tip Ejector	20
7.4.1	Removing the Tip Ejector.....	20
7.4.2	Installing the Tip Ejector.....	21

7.5	Cleaning and Inspecting the Pipette.....	22
7.5.1	Disassembling the Bottom Section of the Pipette (Single-Channel Pipettes).....	22
7.5.2	Cleaning and Greasing the Bottom Section of the Pipette (Single-Channel Pipettes).....	23
7.5.3	Assembling the Lower Part of the Pipette (Single-channel Pipettes).....	24
7.6	Sterilizing the Pipette.....	24
7.6.1	Sterilizing the Pipette Using UV Radiation.....	24
7.7	Defining the Test Routine for the Routine Test.....	25
7.8	Performing the Routine Test.....	26
7.8.1	Aspirating the Test Water.....	26
7.8.2	Dispensing Test Water and Performing the Measurement.....	27
7.8.3	Evaluating Measured Values.....	27
7.9	Adjusting the Pipette.....	28
8	Malfunctions.....	29
9	Storage and Shipping.....	30
9.1	Storing.....	30
9.2	Returning the Device and Parts.....	30
10	Disposal.....	31
10.1	Disposing of the Device and Parts.....	31
11	Technical Data.....	32
11.1	Dimensions and Weights.....	32
11.1.1	Single-Channel Models with Variable Volume.....	32
11.1.2	Multichannel Models with Variable Volume.....	32
11.1.3	Single-Channel Model with Fixed Volume.....	32
11.2	Ambient Conditions at the Installation Site.....	33
11.3	Materials.....	34
11.3.1	Upper Part and Lower Part of Housing.....	34
11.3.2	Model-Dependant Materials of Lower Part of Housing (lower part) and Tip Ejector Components.....	35
11.4	Cleaning Procedure and Cleaning Agent.....	36
11.5	Adjustment.....	37
11.5.1	Initial Adjustment.....	37
11.6	Routine test.....	37
11.6.1	Conditions for Aspirating the Test Liquid.....	37
11.6.2	Correction Factor Z ($\mu\text{L}/\text{mg}$) for Routine Test.....	37
11.7	Performance Specifications.....	38
11.7.1	Single-Channel Models with Variable Volume.....	38
11.7.2	Multichannel Models with Variable Volume.....	40
11.7.3	Single-Channel Model with Fixed Volume.....	41
11.8	Conformity.....	41
12	Accessories and Consumables.....	42
12.1	Accessories.....	42
12.2	Consumables.....	42
13	Sartorius Service.....	43

1 About These Instructions

1.1 Validity

These instructions are part of the device; they must be read in full and kept in a safe place. These instructions apply to the device in the following versions:

Device	Volume Range (µl)	Maximum Volume / Nominal Volume (µl)	Item number
Proline® Mechanical Pipette, with variable volume			
Single-channel	0.25 – 2.5	2.5	720010
	1 – 10	10	720015
	2 – 20	20	720080
	5 – 50	50	720025
	10 – 100	100	720050
	20 – 200	200	720070
	100 – 1000	1000	720060
	1000 – 5000	5000	720110
8-channel	1 – 10	10	720210
	5 – 50	50	720220
	50 – 300	300	720240
12-channel	1 – 10	10	720310
	5 – 50	50	720320
	50 – 300	300	720340
Proline® Mechanical Pipette, with fixed volume			
Single-channel	5	5	722001
	10	10	722004
	20	20	722010
	25	25	722015
	50	50	722020
	100	100	722025
	200	200	722030
	250	250	722035
	500	500	722040
	1000	1000	722045

Device	Volume Range (µl)	Maximum Volume / Nominal Volume (µl)	Item number
Single-channel	2000	2000	722050
	5000	5000	722055

1.2 Related Documents

- In addition to these instructions, observe the following documents: Instructions for consumables, accessories, e.g., pipette holder, pipette tips

1.3 Target Groups

These instructions are addressed to the following target groups. The target groups must possess the specified knowledge.

Target Group	Knowledge and Qualifications
Operator	The operator is familiar with the device and the associated work processes. The operator understands the hazards which may arise when working with the device and knows how to prevent them.

1.4 Symbols Used

1.4.1 Warnings in Operation Descriptions

NOTICE

Denotes a hazard that may result in property damage if it is **not** avoided.

1.4.2 Other Symbols Used

- Required action: Describes activities that must be carried out. The activities in the sequence must be carried out in succession.
- ▷ Result: Describes the result of the activities carried out.

2 Safety Instructions

2.1 Intended Use

The Proline® pipette is a universal laboratory device. The pipette is an air-cushion piston pipette and is intended for dispensing liquids in a variety of applications.

The device's volume range must be taken into consideration when selecting the application (see Chapter "11.7 Performance Specifications", page 38). Volume deviations can be caused by dosing liquids with different physical properties.

Pipette tips suitable for the device must be used (see Chapter "12.2 Consumables", page 42). Using suitable tips guarantees the pipette's performance specifications and ensures pipetting accuracy and precision.

The tips are designed for single use and need to be disposed of after the single use.

The device is suitable for left- and right-handed users.

The device is intended solely for use in accordance with these instructions. Any other use is considered **improper**.

Operating Conditions for the Device

Do **not** use the device in potentially explosive environments. Only use the device indoors.

The device may only be used with the equipment and under the operating conditions described in the Technical Data section of these instructions.

2.1.1 Modifications to the Device

If the device is modified: Persons may be put at risk. Device-specific documents and product approvals may lose their validity.

Contact Sartorius if you have any questions concerning modifications.

2.1.2 Repairs to the Device

Device repairs may only be carried out by persons with specialized knowledge of the device. If the device is **not** repaired by a specialist: Persons may be put at risk. Device-specific documents and product approvals may lose their validity.

Sartorius recommends that any repair work, even that carried out after the end of the warranty period, is carried out by Sartorius Service or after consulting with Sartorius Service.

2.2 Qualifications of Personnel

Persons who do not possess adequate knowledge about how to use the device may injure themselves and other persons.

If a particular qualification is required for an activity: The target group is specified. If **no** qualification is specified: The activity can be performed by the "Operator" target group.

2.3 Functionality of the Device Parts

Non-functioning device parts, e.g., as a result of damage or wear, can cause malfunctions. There is a risk of injury to persons.

- ▶ If device parts are **not** functioning: Do **not** use the device.

2.4 Risk of Injury During Prolonged Pipetting

Prolonged pipetting can cause Work Related Upper Limb Disorder (WRULD) and Repetitive Strain Injury (RSI).

- ▶ Take regular breaks during pipetting and in doing so relax the hand.
- ▶ Stand or sit upright during pipetting.
- ▶ Place the index finger under the finger support and operate the device with the thumb.
- ▶ Do **not** cramp during pipetting.

3 Device Description

3.1 Device Overview




Fig. 1: Single-channel and multichannel models (example)

Pos.	Name
1	Operating button
2	Finger support
3	Tip ejector button
4	Handle
5	Display
6	Tip ejector collar
7	Tip cone
8	Housing for tip cone
9	Tip ejector bar
10	Tip

3.2 Pipette Tips

Sartorius tips are made of pure virgin polypropylene and manufactured in protected cleanroom conditions.

Sartorius non-filter tips are available in trays, bulk packages, and space-saving refill systems. Both non-filter tips and trays are autoclavable at 121°C (252°F), 20 min, 1 bar (15 psi). All Single Tray racks and clean Refill Packs are certified RNase, DNase, and endotoxin-free.

3.2.1 Safetyspace® Filter Tips

To avoid aerosol contamination, use Safetyspace® filter tips and discard them after each pipetting. These can also be used to avoid sample loss caused by the sample coming into contact with the filter. The extra air gap between the sample and the filter ensures that even foaming or viscous liquids do not come into contact with the filter in reverse pipetting or repetitive/multiple dispensing modes.




Fig. 2: Safetyspace® filter tip

Pos.	Name
1	Safetyspace® air gap

3.3 Adjustment Tool

The adjustment tool is required to adjust the pipette and perform assembly work on the pipette.




Fig. 3: Adjustment tool

Pos.	Name
1	For removing the tip ejector collar
2	For removing the tip cone

3.4 Pipetting Practice

The Proline® pipette is used for forward and reverse pipetting.

3.5 Initial Adjustment and Follow-up Adjustment

Initial adjustment: The pipette is adjusted on delivery in accordance with ISO 8655-1 (see Chapter “11.5 Adjustment”, page 37).

Follow-up adjustment: The pipette can be re-adjusted, e.g., if it is determined that the delivered volume is **not** within the error limits or a liquid other than water is to be pipetted. The described procedure must be used for the adjustment.

The current pipette adjustment is displayed in the settings window for adjustment.

4 Installation

4.1 Scope of Delivery

Item	Quantity
Proline® Mechanical Pipette	1
Adjustment tool	1
Safe-Cone Filter (not for 2.5 µl 5 µl 10 µl models)	12
Tweezers (not for 2.5 µl 5 µl 10 µl models)	1
Additional tip ejector collar for 200 µl tips; can replace the tip ejector collar installed upon delivery	
Pipette with a variable volume (only for volumes of 2.5 µl 5 µl 10 µl)	1
Pipette with a fixed volume (only for volumes of 5 µl 10 µl)	1
Quality Control Certificate	1
Quick Start Guide	1

4.2 Unpacking

Procedure

- ▶ Unpack the device.
- ▶ Sartorius recommends keeping the original packaging to return the device appropriately, e.g., for repairs.


4.3 Storing the Pipette When Not in Use

For convenience and safety, always keep the pipette upright. The supplied pipette holder or a pipette stand can be used for this purpose.

The pipette holder must be affixed to a suitable surface.

Procedure

- ▶ Clean the shelf surface with ethanol.
- ▶ Remove the protective paper from the adhesive tape.
- ▶ Install the holder by pressing it against the edge of the shelf and hold for 30 seconds.
- ▶ Allow a waiting time of 20 minutes. This allows the adhesive tape and surface to bond.
- ▶ Hook the finger support onto the holder.



5 Commissioning

5.1 Checking the Replacement of the Tip Ejector

If the pipette is equipped with an additional tip ejector, the additional tip ejector can be installed. This enables 200 µl tips to be used.

Procedure

- Check whether the replacement of the tip ejector is preferred and possible (see Chapter „7.4 Replacing the Tip Ejector“, page 20).

5.2 Adjusting the Volume (Only Pipettes with Variable Volumes)

The volume to be aspirated (pipetted volume) is shown on the display. To adjust the volume rotate the operating button.

NOTICE


Damage to the setting mechanism!

The setting mechanism can become damaged if turned too far.

- Ensure that you do not turn the volume setting beyond the specified range.
-

Procedure

- To set the preferred pipetting volume: Turn the operating button clockwise or counterclockwise.
- ▷ The set volume clicks into place.




5.3 Attaching the Tip

Sartorius recommends using suitable pipette tips for maximum accuracy and precision (see Chapter „12.2 Consumables“, page 42).

Procedure

- ▶ Make sure the pipette tip cone(s) are clean and undamaged.
- ▶ Press the tip cone(s) carefully into a tip held in a tip tray.
- ▶ **Do not** force the pipette into the tip, or rock it from side to side, or front to back when attaching tips.
- ▶ Both sealing surfaces must be clean. The tip is fitted securely when the sealing ring is visible between the sealing surfaces.




5.4 Ejecting the Tip

The pipette has a tip ejector, which reduces the risk of contamination.

Procedure

- ▶ Hold pipette over a suitable waste vessel.
- ▶ Press the tip ejector button with your thumb.



5.5 Inserting or Removing a Safe-Cone Filter

Replaceable Sartorius Safe-Cone Filters can be used with your pipette to help prevent liquids and aerosols from entering the pipette housing, and causing contamination or damage.

We recommend to use Safe-Cone filters when using non-filter tips (not for 2.5 µl | 5 µl | 10 µl models). Do **not** use a Safe-Cone filter with a filter tip.

Safe-Cone Filters are available in Standard or Plus versions. The Standard filter is recommended for general applications, the Plus filter for more demanding applications, such as:


- Cell culture
- Bacterial and virological work
- Molecular biology

The Safe-Cone filter must be changed regularly. The change interval depends on the application, but regular daily replacement is recommended and always in the case of over-aspiration. If the pipette is **not** functioning properly, change the filter.

5.5.1 Inserting a Safe-Cone Filter

Procedure

- Clean the tip cone if necessary.
- Insert a new filter.



5.5.2 Removing a Safe-Cone Filter

Procedure

Remove the Safe-Cone Filter with the provided tweezers.

6 Operation

6.1 Considering Good Pipetting Practices


- Make sure the tip is firmly attached to the tip cone(s).
- Pre-rinse the tip before aspirating by filling and emptying the tip three to five times. This is particularly important when pipetting liquids with a viscosity and density greater than water, or volatile liquids with high vapor pressure (such as ethanol).
- Make sure the pipette, tip, and liquid are at the same temperature (if possible).
- Hold the pipette upright when aspirating, and only place the tip a few millimeters into the liquid (recommended immersion depth see Chapter “11.6.1 Conditions for Aspirating the Test Liquid”, page 37).
- Always press and release the operating button slowly and smoothly.
- When pipetting a liquid that is **not** at the ambient temperature, the tip must be changed after each pipetting. Do **not** rinse the tip before use.
- To avoid contaminating the pipette’s housing with either the sample in the tip or the work surface, do **not** lay the pipette on its side.
- Never strike the tip cone against the tip tray.
- When handling infectious or radioactive agents, wear appropriate protective gear and take all sensible precautions.
- Avoid exposing the unit to extreme temperature changes, humidity, or dust. Observe the operating conditions (see Chapter “11.2 Ambient Conditions at the Installation Site”, page 33).

6.2 Forward Pipetting

Forward pipetting is the most common pipetting technique. It uses a blow-out function to ensure complete delivery of the liquid. This technique aspirates and then dispenses the selected liquid volume. It is recommended for aqueous liquids, liquids containing small amounts of detergents or proteins, and solvents.

Procedure

- ▶ Fit the tip onto the pipette tip cone.
- ▶ Press the operating button to the first stop.
- ▶ Place the tip just under the surface of the liquid and smoothly release the operating button, allowing it to return to the starting position. Wait one second.
- ▶ Carefully withdraw the tip from the liquid, touching it against the edge of the container to remove any excess liquid.
- ▶ Dispense the liquid by pressing the operating button to the first stop (1).
- ▶ After a short pause, press the operating button to the second stop (2). This empties the tip.




6.3 Reverse Pipetting

Reverse pipetting is recommended for highly viscous, biological, or foaming liquids, or very small volumes. The selected volume is aspirated into the tip along with an excess volume.

Delivery is done without blow-out ensuring the excess volume remains in the tip. This excess is then discarded after the selected volume is aspirated. You may need to adjust the pipette in order to use the reverse pipetting technique.

Procedure

- ▶ Fit the tip onto the pipette tip cone.
- ▶ Press the operating button all the way to the second stop.
- ▶ Place the tip just under the surface of the liquid and smoothly release the operating button, allowing it to return to the starting position.
- ▶ Carefully withdraw the tip from the liquid, touching it against the edge of the container to remove any excess from the tip.




- ▶ Press the operating button smoothly to the first stop (1) to deliver the preferred volume. The liquid that remains in the tip may **not** be included in the delivery.
- ▶ Discard the remaining liquid by pressing the operating button to the second stop (2).

6.3.1 Repetitive Reverse Pipetting

The reverse technique allows you to repeatedly deliver the same volume.

Procedure

- ▶ Fit the tip onto the pipette tip cone.
- ▶ Press the operating button all the way to the second stop.
- ▶ Repeat the following steps as many times as necessary:
 - ▶ Place the tip just under the surface of the liquid and smoothly release the operating button, allowing it to return to the starting position.
 - ▶ Carefully withdraw the tip from the liquid, touching it against the edge of the container to remove any excess from the tip.
 - ▶ Press the operating button smoothly to the first stop (1) to deliver the preferred volume. The liquid that remains in the tip may **not** be included in the delivery.
- ▶ When finished dispensing: Discard the remaining liquid by pressing the operating button to the second stop (2).



7 Cleaning and Maintenance

7.1 Testing Chemical Resistance

Information on interactions of pipette materials with organic and inorganic solutions, solvents and corrosive chemicals can be provided on request.

Procedure

- Check that the pipette materials are chemically compatible with the detergent, disinfectant, or decontaminant liquids to be used (pipette materials see Chapter “11.3 Materials”, page 34).

7.2 Cleaning the Outer Surface of the Pipette

We recommend to clean the surface of the device daily. To clean and decontaminate the outer surface of the pipette, use a disinfectant liquid or mild detergent and a soft, lint-free cloth.

The cleaning agent and decontamination agent must be suitable for the device materials (see Chapter “11.3 Materials”, page 34 and Chapter “11.4 Cleaning Procedure and Cleaning Agent”, page 36).

We recommend using gloves when cleaning the pipette.

Procedure

- Daily check the Proline® pipette to ensure it is clean.
- Gently clean the surface of the pipette with the moistened cloth and wipe it dry. Pay special attention to the tip cone.
- Replace the Safe-Cone filter.

7.3 Maintenance Schedule

Interval	Component	Activity	Chapter, page
Regularly, depending on the operating conditions	Safe-Cone Filter	Replace the Safe-Cone Filters.	5.5, 15
	Single-channel models: Lower part of the device	Clean the lower parts and grease components.	7.5, 22
	Multi-channel models: Lower part of the device	For cleaning and greasing the lower parts contact Sartorius Service.	
	Device	Sterilise the device.	7.6, 24
Regularly, e.g., every 3 months and after each in-house maintenance	Device	Perform a routine test to check the pipette performance.	7.8, 26
6 to 24 months, depending on the operating conditions	Device	Contact Sartorius Service for calibrating the device.	

7.4 Replacing the Tip Ejector

If the pipette is equipped with an additional tip ejector, the additional tip ejector can be installed. This enables 200 µl tips to be used.


Procedure


- Check whether the pipette is fitted with a tip ejector (see Chapter “4.1 Scope of Delivery”, page 12.).

7.4.1 Removing the Tip Ejector

Procedure

- Press and hold down the tip ejector button.
- Use the adjustment tool to lift the engaged lock between the tip ejector sleeve and the tip ejector.






- Hold down the tip ejector button and pull the tip ejector off.

7.4.2 Installing the Tip Ejector

Procedure



- Install the tip ejector. To do so, complete the following steps:
 - Hold down the tip ejector button.
 - Allow the notch of the tip ejector sleeve to engage in the hole in the tip ejector.
- Release the tip ejector button.
- Check whether the tip ejector is working correctly by pressing the tip ejector button again.


7.5 Cleaning and Inspecting the Pipette

7.5.1 Disassembling the Bottom Section of the Pipette (Single-Channel Pipettes)

The information below only describes single-channel pipettes. If you need to clean multichannel pipettes, contact Sartorius Service.

Procedure

- ▶ If available: Remove the tip cone filter (see Chapter “5.5 Inserting or Removing a Safe-Cone Filter”, page 15).
- ▶ Remove the tip ejector (see Chapter “7.4.1 Removing the Tip Ejector”, page 20).
- ▶ Depending on the pipetting volume, disassemble the bottom section of the pipette as follows:
 - ▶ Models with volumes of 2.5 µl to 5000 µl: Loosen the tip cone cylinder with the adjustment tool (not an external tool) by turning counterclockwise.
 - ▶ Models with volumes of 5000 µl: Turn the tip cone cylinder counterclockwise without using a tool.
- ▶ Carefully remove the tip cone cylinder, the piston and the spring.



7.5.2 Cleaning and Greasing the Bottom Section of the Pipette (Single-Channel Pipettes)

Material: Silicone grease

Avoid excess grease, too much grease can block the piston. Only use grease that is recommended by Sartorius.

The following information only applies to single-channel pipettes. For cleaning multi-channel pipettes contact Sartorius Service.

Piston seal: For the following models, the piston seal cannot be accessed for cleaning:

- Pipette with a variable volume of: 0.25 µl–2.5 µl | 1 µl–10 µl
- Pipette with fixed volume: 5 µl | 10 µl

Procedure

- ▶ Clean the tip ejector, tip cone cylinder, piston and piston seal with a disinfectant solution or a mild detergent and a soft, lint-free cloth.
- ▶ Clean the inside of the tip ejector and the tip cone cylinder with a cotton swab. Be extremely careful with the following models, in order to prevent damage to the seal in the tip cone: Models with volumes of 2.5 | 5 | 10 | 20 | 25 | 50 | 100 µl.
- ▶ If necessary, clean the sections with distilled water and allow them to dry.
- ▶ Apply a thin layer of silicone grease to the pistons or the piston seal.
- ▶ Only models with volumes from 20 µl: Apply a thin layer of silicone grease to the piston seal.


7.5.3 Assembling the Lower Part of the Pipette (Single-channel Pipettes)

Before reassembling, check that there is **no** lint or other particles on the surface of the piston.

The following information only applies to single-channel pipettes. For cleaning multi-channel pipettes contact Sartorius Service.

Procedure

- ▶ Depending on the pipette volume reassemble the lower parts as follows:
 - ▶ Place the spring and tip cone cylinder on the piston.
 - ▶ Models with volumes of 2.5 µl to 5000 µl: Tighten the tip cone cylinder with the adjustment tool (not an external tool) by turning clockwise. Do not screw it on too tightly.



- ▶ Models with volumes of 5000 µl: Tighten the tip cone cylinder without a tool by turning clockwise.
- ▶ Attach the tip ejector collar by screwing it clockwise.
- ▶ Insert a new tip cone filter.
- ▶ Press the operating button several times to ensure that the grease has spread evenly.
- ▶ Perform a routine test for the pipette.

7.6 Sterilizing the Pipette

The Proline® Pipette can be sterilized by means of UV light radiation and a disinfectant or using a decontamination liquid. Always follow the instructions below.

7.6.1 Sterilizing the Pipette Using UV Radiation

Proline® pipettes are made of UV-resistant materials and tolerate temporary exposure to UV radiation. Take note that prolonged or frequent exposure to UV radiation may cause yellowing and brittling of the pipette.

7.7 Defining the Test Routine for the Routine Test

We recommend testing the performance of the pipette regularly in a routine test (for intervals, see Chapter “7.3 Maintenance Schedule”, page 20).

We recommend defining a test routine for the routine test that takes into account the following criteria:

Test Routine	Criteria	Explanation
General test routine	Accuracy requirements of the relevant application	
	Frequency of use	
	Number of operators for the device	
	Nature of the liquid dispensed	
	Maximum permitted error limits for systematic and random errors, as per ISO 8655-2	As per ISO 8655-2
	Acceptable error range taking into account: – Application – Application area – Conditions for precision of the application	Sartorius specifications were achieved in strictly controlled conditions, as per ISO 8655-6 (see Chapter “11.7 Performance Specifications”, page 38).
	Multistage routine test with three volumes recommended: – 100% of the nominal volume – 50% of the nominal volume – 10% of the nominal volume	
	Risk of the application, e.g. in relation to accuracy	
Additional requirements for multi-channel pipettes	All channels of the multi-channel pipette must be tested individually. The following options are available for doing this:	
	Use a multi-channel balance that measures the dispensed liquid from all channels in parallel.	The test liquid is aspirated, dispensed and measured in all channels at the same time.
	Use a single-channel balance that measures the dispensed test liquid from one channel.	The test liquid for each channel is aspirated, dispensed and measured individually in succession. The test liquid from the other channels is discarded as residual water.

Procedure

- Prepare a test routine that takes into account the required criteria for the routine test on the device.

7.8 Performing the Routine Test

A routine test with an analytical balance (single-channel balance) is described below. Further information on the possible test routines is described in the ISO 8655 series of standards, e.g., for the testing of multi dispensing in pipetting mode.

- Materials:
- Test liquid (test water): distilled, deionized water (ISO 3696, quality 3)
 - Test water container
 - Measuring container, for positioning on the analytical balance
 - Pipette tips
- Tool: Analytical balance, in compliance with the standard ISO 8655-6

Requirements

The test water container is filled with test water.

Procedure

- ▶ Check whether the following conditions are met:
 - The required ambient conditions have been complied with (see Chapter “11.2 Ambient Conditions at the Installation Site”, page 33).
 - The device, tips and test water have been stored in the room for at least two hours so that they have adjusted to the ambient conditions.
 - The “Pipetting” pipetting mode is set in the device.
 - The aspirating speed and dispensing speed are set to 7.
- ▶ If necessary: Ensure that the conditions are complied with or adjust the settings on the device.

7.8.1 Aspirating the Test Water

The conditions for aspirating the test water must be complied with (see Chapter “11.6.1 Conditions for Aspirating the Test Liquid”, page 37).

Procedure

- ▶ Adjust the preferred test volume (V_S).
- ▶ Apply the tip to the tip cone.
- ▶ To achieve a humidity balance in the device: Aspirate and dispense the test water 5 times with the tip.
- ▶ Aspirate the test water. To do this, proceed as follows:
 - ▶ Hold the device vertically, press the operating button and immerse the tip below the surface of the water to the specified immersion depth.
 - ▶ Slowly release the operating button.
 - ▶ Comply with the specified wait time.
- ▶ Remove the device from the test water container vertically.

7.8.2 Dispensing Test Water and Performing the Measurement

We recommend performing 10 measurements. The tip must be replaced after 5 measurements.

Requirements

The analytical balance is ready for measuring.

Procedure

- ▶ Dispense the test water. To do this, proceed as follows:
 - ▶ Hold the tip above the liquid level at an angle of 30° – 40° on the inner wall of the measuring container.
 - ▶ Press and hold the operating button and release the test water into the measuring container.
 - ▶ In order to remove drops on the tip: Drag the tip 8 – 10 mm along the inner wall of the measuring container.
- ▶ Release the operating button.
- ▶ Remove the device from the measuring container vertically.
- ▶ Read off the weight in mg (m_i).
- ▶ Repeat the test cycle until 10 measurements have been recorded. Replace the tip after five test cycles.

7.8.3 Evaluating Measured Values

Requirements

Ten measurements have been recorded.

Procedure

- ▶ Convert the recorded masses (m_i) into volume (V_i) by multiplying with a correction factor Z. The following applies here: $V_i = m_i \cdot Z$ (for correction factor Z, see Chapter 11.6.2, page 37).
- ▶ Calculate the average volume: (V): $V = \Sigma(V_i)/10$.
- ▶ To evaluate conformity: Calculate the systematic error " e_s " of the measurement. Use the following formula to do this:
 - In μL : $e_s = V - V_s$
 - Or in %: $e_s = 100 (V - V_s)/V_s$
 - V_s = selected test volume
- ▶ For the conformity assessment, calculate the random measurement deviation as standard deviation or as variation coefficient:
 - As standard deviation (n = number of measurements (10))
 - As variation coefficient $C_v = 100\% \text{ sr} / V$

$$s = \sqrt{\frac{\Sigma(V_i - \bar{V})^2}{n - 1}}$$


- ▶ Compare the systematic error and the random error with the values of the performance specification from your own laboratory.
- ▷ If the measurement results are within the performance specification: The device is ready for use.
- ▷ If the measurement results are **not** within the performance specification:
 - ▶ Check the device for both systematic errors and random errors.
 - ▶ Ensure that a suitable tip is used.
 - ▶ Ensure that a good seal has been formed between tip and tip cone.
- ▷ If necessary: Adjust the device (see Chapter 7.9, page 28).

7.9 Adjusting the Pipette

The pipette is adjusted on delivery. A follow-up adjustment is required under the following conditions:

- When pipetting liquids which differ from water in density, viscosity, vapor pressure, or other quality.
- When reverse pipetting.
- When the geometry of the tip used clearly differs from standard tip geometry.
- When the pipette is used at high altitudes.
- The pipette routine test indicates a performance specification error of the pipette.

Procedure



- ▶ Insert the adjustment tool supplied into the calibration notch below the operating button.
- ▶ Turn the adjustment tool counterclockwise to reduce the volume or clockwise to increase the volume.
- ▶ Repeat checks of the pipette until the test results are correct (see Chapter "7.8 Performing the Routine Test", page 26).
- ▶ If the pipette has been adjusted for the use of liquids other than water, attach a label to the outside of the pipette with the name of the liquid and the adjusted volume range. To do this, the supplied stickers can be used for labeling.

8 Malfunctions

Problem	Possible Cause	Solution
Droplets left in the tip	Unsuitable tip	Only use original Sartorius tips
Leakage / pipetted volume too small	Pre-rinsing the tip was not successful	Pre-rinse the tip again
	Tip incorrectly attached	Attach tip securely
	Unsuitable tip	Only use original Sartorius tips
	Foreign particles between the tip and the cone	Clean tip cone, attach new tip
	The device has a leak	Clean and grease the piston seal
	Tip cone holder incorrectly tightened	Tighten tip cone holder
	Pipette damaged	Send to Sartorius Service for servicing
Pipette does not comply with the established test specifications	Incorrect operation	Refer to the instructions
	Unsuitable tip	Only use original Sartorius tips
	Adjustment setting altered	Readjust the pipette
Operating button jammed or moves erratically	Liquid has penetrated the tip cone and dried	Clean and grease piston and seal, clean tip cone
	Solvent vapors have penetrated into the tip cone	
	The piston is contaminated.	
	Safe-Cone Filter has been contaminated	Change filter
	Insufficient amount of grease on piston and seal	Grease as required
Tip ejector jams or moves erratically	Tip ejector has been contaminated	Remove and clean tip ejector and tip cone

9 Storage and Shipping

9.1 Storing

Procedure

- ▶ Clean the device.
- ▶ Store the device according to the ambient conditions (see Chapter “11.2 Ambient Conditions at the Installation Site”, page 33).
- ▶ Store the device vertically, e.g., in a pipette stand.

9.2 Returning the Device and Parts

Defective devices or parts can be returned to Sartorius. Returned devices must be clean and properly packed.

Transport damage as well as measures for subsequent cleaning and disinfection of the device or parts by Sartorius are charged to the sender.

Devices contaminated with hazardous materials, e.g., harmful biological or chemical substances, will **not** be accepted for repair or disposal.

Procedure

- ▶ Ensure that the pipette has been decontaminated before you send it for repair or calibration (see Chapter “7.6 Sterilizing the Pipette”, page 24). Be sure to note any hazardous material that the pipette has been exposed to.
- ▶ Contact Sartorius Service for instructions on how to return devices or parts (see www.sartorius.com).
- ▶ Pack the device and its parts properly for return.

10 Disposal

10.1 Disposing of the Device and Parts

The device and the device accessories must be disposed of properly by disposal facilities.

Requirements

The device is decontaminated.

Procedure

- ▶ Dispose of the device in accordance with local government regulations.
- ▶ Dispose of the packaging in accordance with local government regulations.
- ▶ Dispose of the consumables in accordance with local government regulations.

11 Technical Data

11.1 Dimensions and Weights

11.1.1 Single-Channel Models with Variable Volume

		720010	720015	720080	720025	720050	720070	720060	720110
	Unit	Value	Value	Value	Value	Value	Value	Value	Value
Dimensions (length)	mm	257	255	232	241	224	232	232	202
Weight	g	75	75	75	75	80	75	85	90

11.1.2 Multichannel Models with Variable Volume

		720210	720220	720240	720310	720320	720340
	Unit	Value	Value	Value	Value	Value	Value
Dimensions (length)	mm	202	250	250	250	250	250
Weight	g	150	155	165	175	190	200

11.1.3 Single-Channel Model with Fixed Volume

Models with Volumes of 5 µl to 250 µl

		722001	722004	722010	722015	722020	722025	722030	722035
	Unit	Value	Value	Value	Value	Value	Value	Value	Value
Dimensions (length)	mm	255	255	232	232	232	232	232	232
Weight	g	75	75	75	75	75	75	75	85

Models with Volumes of 500 µl to 5000 µl

		722040	722045	722050	722055
	Unit	Value	Value	Value	Value
Dimensions (length)	mm	232	232	202	202
Weight	g	85	85	90	90

11.2 Ambient Conditions at the Installation Site

	Unit	Value
Laboratory, for indoor use only		
Pollution level according to IEC 61010-1		II
Maximal height above sea level	m	2000
Suitable for protection class		
Protection class of the device, as per DIN EN 60529-1		IP 40
Temperature		
In operation	°C	+15 – +30
During performance review	°C	+15 – +30
In storage during transport	°C	-20 – +40
Relative humidity, in operation		
At temperatures up to 31°C, maximum	%	80
No potentially explosive areas		
Other properties		
Store dry		

11.3 Materials

11.3.1 Upper Part and Lower Part of Housing

Component	Materials										
	Polyetherimide (PEI)	Polyethylene terephthalate (PETP)	Acrylonitrile butadiene styrene copolymers (ABS)	Polypropylene (PP)	Polyamide (PA)	Polycarbonate (PC)	Polyethylene (PE)	Thermoplastic polyurethane (TPU)	Polyvinylidene fluoride (PVDF)	Acrylonitrile styrene acrylate (ASA)	Polyethylene terephthalate (PETP)
Top section of the housing											
Operating button	-	-	x	-	x	-	-	-	-	-	-
Finger support	-	-	-	-	-	x	-	-	-	x	-
Tip ejector button	-	-	-	-	x	-	-	-	-	-	-
Handle	-	-	-	-	-	x	-	-	-	x	-
Display	-	-	-	-	-	x	-	-	-	-	-
Bottom section of the housing (single-channel models)											
Tip ejector (single-channel models)	-	-	-	-	x	-	-	-	-	-	-
Bottom section of the housing (multichannel models)											
Housing for tip cone	-	-	-	-	-	x	-	-	-	x	-
Tip ejector bar	-	x	-	-	-	-	-	-	-	-	-
Tip	-	-	-	-	-	-	-	x	x	-	-
Tip cone	-	-	-	-	-	x	-	-	-	-	-

11.3.2 Model-Dependant Materials of Lower Part of Housing (lower part) and Tip Ejector Components

Component	Materials											
	Polyetherimide (PEI)	Silicone (SI)	Stainless steel (SS)	Polypropylene (PP)	Polyamide (PA)	Polycarbonate (PC)	Polyethylene (PE)	Polyphenylenesulphide (PPS)	Polyvinylidene fluoride (PVDF)	Fluoroelastomer (FKM)	Ethylene propylene diene mono-mer (EPDM)	
Tip cone												
For 2.5 µl 10 µl 20 µl 25 µl 50 µl 100 µl models	-	-	-	-	-	-	-	-	x	-	-	
For 200 µl models, single-channel	-	-	-	-	-	-	-	x	-	-	-	
For 300 µl 500 µl models	x	-	-	-	-	-	-	-	-	-	-	
For 1,000 µl 5,000 µl 10,000 µl models	-	-	-	-	-	-	-	x	-	-	-	
Piston												
For 3 µl 10 µl 20 µl 25 µl 50 µl models	-	-	x	-	-	-	-	-	-	-	-	
For 100 µl single-channel model	-	-	-	-	-	-	-	x	-	-	-	
For 100 µl multi-channel model	x	-	-	-	-	-	-	-	-	-	-	
For 200 µl 300 µl models	x	-	-	-	-	-	-	-	-	-	-	
For 1,000 µl 5,000 µl 10,000 µl models	-	-	-	-	-	-	-	x	-	-	-	
Piston seal												
For 2.5 µl 10 µl 20 µl 10,000 µl models	-	-	-	-	-	-	-	-	-	x	-	
For 100 µl 200 µl 300 µl 500 µl 1,000 µl 5,000 µl models	-	-	-	-	-	-	-	-	-	-	x	

11.4 Cleaning Procedure and Cleaning Agent

	Unit	Value
Approved cleaning procedures		
Wiping the device surfaces with a lightly moistened cloth		
Wiping the device surfaces dry		
UV radiation		
Approved cleaning agents		
Ethanol (70%)		
Isopropyl alcohol (60%)		
Mild cleaning agent, suitable for the device materials		

11.5 Adjustment

11.5.1 Initial Adjustment

Type of adjustment by manufacturer in accordance with ISO 8655-1

On dispensing (Ex) of the selected volume (test volume), at 20 °C

11.6 Routine test

11.6.1 Conditions for Aspirating the Test Liquid

Test volume (μL)	Immersion depth of the pipette tip during aspiration (mm)	Wait time (s)
≤ 1	1 – 2	1
> 1 to 100	2 – 3	1
> 100 to 1 000	2 – 4	1
> 1 000 to 20 000	3 – 6	3

11.6.2 Correction Factor Z (μL /mg) for Routine Test

Temperature (°C)	Air Pressure (kPa)			
	95	100	101.3	105
20.0	1.0028	1.0028	1.0029	1.0029
20.5	1.0029	1.0029	1.0030	1.0030
21.0	1.0030	1.0031	1.0031	1.0031
21.5	1.0031	1.0032	1.0032	1.0032
22.0	1.0032	1.0033	1.0033	1.0033
22.5	1.0033	1.0034	1.0034	1.0034
23.0	1.0034	1.0035	1.0035	1.0036
23.5	1.0036	1.0036	1.0036	1.0037

The values in the table represent the most common values. For the full table, or for formula to calculate correction factor the standard ISO 8655-6 must be considered.

11.7 Performance Specifications

11.7.1 Single-Channel Models with Variable Volume

Model	Color of the operating button	Volume range	Maximum permitted error limits as per ISO 8655				
			Test volume	Systematic errors*		Random errors*	
			Unit	Unit	Unit	Unit	Unit
			µL	µL	%	%	µL
			Value	Value	Value	Value	Value
720010	⑤	0.25–2.5	2.5	3.0	0.075	2.4	0.06
			1.25	5.0	0.0625	4.0	0.05
			0.25	18.0	0.045	14.0	0.035
720015	⑤	1–10	10	1.2	0.12	0.8	0.08
			5	2.0	0.1	1.5	0.075
			1	6.0	0.06	5.0	0.05
720080	⑤	2–20	20	1.0	0.2	0.5	0.1
			10	1.7	0.17	0.9	0.09
			2	7.5	0.15	4.0	0.08
720025	⑤	5–50	50	1.0	0.5	0.4	0.2
			25	1.8	0.45	0.7	0.175
			5	8.0	0.4	3.0	0.15
720050	⑤	10–100	100	0.8	0.8	0.3	0.3
			50	1.2	0.6	0.5	0.25
			10	4.5	0.45	2.0	0.2
720070	⑤	20–200	200	0.8	1.6	0.3	0.6
			100	1.2	1.2	0.4	0.4
			20	4.0	0.8	1.3	0.26

* The values specified for the systematic and random error limits apply under the following conditions:

- When using the specified models with suitable non-sterile Sartorius Optifit tips
- When calculating the values under strictly controlled conditions during type examinations in accordance with ISO 8655

The values specified in ISO 8655 for the systematic and random error limits apply under the following conditions:

- When using the specified models with other Sartorius tips

Model	Color of the operating button	Volume range	Maximum permitted error limits as per ISO 8655				
			Test volume	Systematic errors*		Random errors*	
			Unit	Unit	Unit	Unit	Unit
			µL	µL	%	%	µL
			Value	Value	Value	Value	Value
720060	⑤	100–1000	1000	0.8	8	0.3	3
			500	1.2	6	0.4	2
			100	4.0	4	1.2	1.2
720110	⑤	1000–5000	5000	0.8	40	0.3	15
			2500	1.0	25	0.4	10
			1000	2.0	20	0.7	7

* The values specified for the systematic and random error limits apply under the following conditions:

- When using the specified models with suitable non-sterile Sartorius Optifit tips
- When calculating the values under strictly controlled conditions during type examinations in accordance with ISO 8655

The values specified in ISO 8655 for the systematic and random error limits apply under the following conditions:

- When using the specified models with other Sartorius tips

11.7.2 Multichannel Models with Variable Volume

Model	Number of channels	Color of the operating button	Volume range	Maximum permitted error limits as per ISO 8655					
				Test volume	Systematic errors*		Random errors*		
				Unit	Unit	Unit	Unit	Unit	
				µL	µL	%	µL	%	µL
				Value	Value	Value	Value	Value	
720210 720310	8 12	⑤	1–10	10	2.4	0.24	1.6	0.16	
				5	4.0	0.2	2.8	0.14	
				1	12.0	0.12	8.0	0.08	
720220 720320	8 12	⑤	5–50	50	1.5	0.75	0.7	0.35	
				25	2.5	0.625	1.2	0.3	
				5	10.0	0.5	5.0	0.25	
720240 720340	8 12	⑤	5–300	300	0.9	2.7	0.35	1.05	
				150	1.3	1.95	0.6	0.9	
				50	2.8	1.4	1.4	0.7	

* The values specified for the systematic and random error limits apply under the following conditions:

- When using the specified models with suitable non-sterile Sartorius Optifit tips
- When calculating the values under strictly controlled conditions during type examinations in accordance with ISO 8655

The values specified in ISO 8655 for the systematic and random error limits apply under the following conditions:

- When using the specified models with other Sartorius tips

11.7.3 Single-Channel Model with Fixed Volume

Model	Color of the operating button	Volume range	Maximum permitted error limits as per ISO 8655				
			Test volume	Systematic errors*		Random errors*	
			Unit	Unit	Unit	Unit	Unit
			µL	µL	%	µL	µL
			Value	Value	Value	Value	Value
722001	⑤	5	5	1.3	0.065	1.2	0.060
722004	⑤	10	10	0.8	0.080	0.8	0.080
722010	⑤	20	20	0.6	0.120	0.5	0.100
722015	⑤	25	25	0.5	0.125	0.3	0.075
722020	⑤	50	50	0.5	0.250	0.3	0.150
722025	⑤	100	100	0.5	0.50	0.3	0.30
722030	⑤	200	200	0.4	0.80	0.2	0.40
722035	⑤	250	250	0.4	1.00	0.2	0.50
722040	⑤	500	500	0.3	1.50	0.2	1.00
722045	⑤	1000	1000	0.3	3.0	0.2	2.0
722050	⑤	2000	2000	0.3	6.0	0.15	3.0
722055	⑤	5000	5000	0.3	15	0.15	7.5

* The values specified for the systematic and random error limits apply under the following conditions:

- When using the specified models with suitable non-sterile Sartorius Optifit tips
- When calculating the values under strictly controlled conditions during type examinations in accordance with ISO 8655

The values specified in ISO 8655 for the systematic and random error limits apply under the following conditions:

- When using the specified models with other Sartorius tips

11.8 Conformity

Development and manufacture of the device in accordance with the requirements of the following standards

ISO 9001

ISO 13485

12 Accessories and Consumables

12.1 Accessories

This table contains a selection of the accessories that can be ordered. For information on other products, contact Sartorius.

Item	Quantity	Order number
Pipette holder	1	LH-727640
Carousel stand for 6 pipettes	1	LH-725630
Linear pipette stand	1	725620

12.2 Consumables

This table contains a selection of the consumables that can be ordered. For information on other products, contact Sartorius.

Item	Quantity	Order number
Optifit Tips, 0.1–10 µL, single rack	1	790010
Optifit Tips, 0.1–10 µL, refill pack	1	790013
Optifit Tips, 0.1–10 µL, refill tower	1	790011
Safetyspace® Filter-Tip, 0.1–10 µL, single rack	1	790011F
Safetyspace® Filter-Tip, 0.5–200 µL, single rack	1	790101F
Safetyspace® Filter-Tip, 50–1,000 µL, single rack	1	791001F
Extended Standard Tip, 0.1–10 µL, single rack	1	783210
Extended Standard Tip, 10–1,000 µL, single rack	1	LH-X781000
Extended Filter Tips 0.1–10 µL, single rack	1	783201
Extended Filter Tips 10–1,000 µL, single rack	1	LH-XF781001
Safe-Cone Filter, standard, Ø6.73 mm	50	721005
Safe-Cone Filter, standard, Ø5.33 mm	50	721006
Safe-Cone Filter, standard, Ø3.15 mm	50	721007
Safe-Cone Filter, standard, Ø2.51 mm	50	721008
Safe-Cone Filter, standard, Ø1.83 mm	50	721014
Tweezers for filter replacement	1	721009

13 Sartorius Service

Sartorius Service is available for queries regarding the device. Visit the Sartorius website (www.sartorius.com) for information about the service addresses, services provided, or to contact a local representative.

For inquiries about the device or when contacting Sartorius Service in the event of a malfunction, ensure that you have the device information, e.g., serial number, close at hand so you can pass it on to Sartorius Service. To do this, refer to the information on the manufacturer's ID label.

Sartorius Liquid Handling oy
Tulppatie 1
00880 Helsinki, Finland

Lhinfo.finland@sartorius.com
www.sartorius.com

The information and figures contained in these instructions correspond to the version date specified below.

Sartorius reserves the right to make changes to the technology, features, specifications and design of the equipment without notice.

Masculine or feminine forms are used to facilitate legibility in these instructions and always simultaneously denote all genders.

Copyright notice:

These instructions, including all components, are protected by copyright.

Any use beyond the limits of the copyright law is not permitted without our approval.

This applies in particular to reprinting, translation and editing irrespective of the type of media used.

Last updated:

03 | 2025

© 2025

Sartorius Liquid Handling oy
Tulppatie 1
00880 Helsinki, Finland

MV | Publication No.: WPR6007-e250303