Specification

1.09967.0001 Aluminium standard 1000 mg Al, ($AlCl_3$ in H_2O) Titrisol®

	Specification
Assay (Al)	1000 mg / Ampoule

Determination method: ICP - OES
(traceable to NIST - SRM 3101a)
Accuracy of the method: ± 2 mg

Ayfer Yildirim
Responsible laboratory manager quality control

This document has been produced electronically and is valid without a signature.