


CERTIFICATE OF CONFORMANCE

Catalog #:	15CTAPP-CF-DSL-ST		
Description:	OmniTop Sampling Assembly 15mL Polypropylene Sampling Tube with 18 of C-Flex Tubing and Tygon Seal Line, with extended diptube.		
Lot Number:	21118-1806		
Client PO#:	4511726102/E40		
Quantity:	400		
DOM:	Nov 2018	Expiration Date/Warranty: Nov 2021	
Gamma ID	2352751		

- OmniTop Sample Tubes® are manufactured and packaged in a certified ISO 7 Class Clean Room conforming to the following standards:
 - ISO 14644-1:1999
- OmniTop Sample Tubes® are manufactured to meet or exceed all design specifications. Performance is confirmed by testing as required or on an audit basis.
- OmniTop Sample Tubes® come with a 3-year manufacturer's warranty against defects in workmanship or materials, post-irradiation. Refer to your product label for manufacturing & expiration dates.
- All fluid contact surfaces meet the criteria for the USP <88> Class VI Biological Test for Plastics.
- All raw materials and processes used in the fluid path of this product are free from any animal derived materials.



R. Scott Jennings
Quality Assurance



38 Industrial Way East, #5
Eatontown, NJ 07724
Phone 732-460-0088, fax 732-460-0802
www.pawbio.com