

Corning Incorporated - Life Sciences
 Antigua Carretera A Roma KM 3.6
 San Nicolas De Los Garza NL 66480 MEX
www.corning.com/lifesciences
 Refer to website for regional contact information.

Page: 1 / 1

Product Name : 5642 FLASK 100ML	Manufacture Date : 2025-11-04
Catalog Number : 5642-100	
Lot ID : 30825084	
Expiration Date : Unconfirmed Validity	

Quality Management System - Complies with the current version of the ISO 9001 Standard.

Glass Type - Borosilicate, Type I, Class A, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

Volumetric Tolerance - Volumetric Accuracy at fill line +/- 0.08 mL.

Product Specification Standards - Product made in accordance with:

- ASTM E288 - Standard Specification for Laboratory Glass Volumetric Flasks
- ASTM E542 - Specification for Calibration of Laboratory Volumetric Apparatus
- ASTM E694 - Standard Specification for Laboratory Glass Volumetric Apparatus,

Class A Criteria

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

- Visual Inspection - Pass
- Packaging Inspection - Pass
- Dimensional Test - Pass
- Capacity Verification - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.


 Ricardo Carrizalez
 Quality Manager