



Certificate of Conformity

VWR Precision Thermometer

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

Cat. No.	Lot	Range (°C)	Accuracy (°C)	Division (°C)	Length (mm)	Pk
620-0807	408546	-1...+101	±0.6	0.1	610	1

Material: Blue Organic

Other:

Statement of Accuracy and Traceability to NIST

Quality Conformance Testing

The above product was manufactured and tested against thermometer standards traceable to the National Institute of Standards and Technology (NIST). The Standard Serial Number is based on the range of the thermometer. The Standard Serial Numbers calibrated by NIST are as follows.

#844016 range -30°C to 10°C

#878708 range 0°C to 50°C

This document has been produced electronically and is valid without a signature

Date of issue: **November 2020**