

Certificate of Conformity

VWR Disposable Triangle Shaped Spreader, Sterile

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Cat. No.	Lot No.	Width	Colour	Packed	Pk
612-2688	042023	30 mm	Blue	Individually Wrapped	25

Date of Sterilization: 15th May 2023

Sterilization Expiry date: 5 years after sterilization date unless otherwise compromised.

Dosage Range: 25-25 kGy

Sterilization Process: E-beam

Material: Virgin Polypropylene

Quality System Compliance: Products are manufactured under GMP guideline standards. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

This product has passed requirements for visual inspection, dimensional and/or physical property inspection.

BSE/TSE: Product is free from BSE/TSE

RNASE/DNASE/PYROGEN: Product is tested and proven RNASE/DNASE/PYROGEN free.

Cytotoxicity: Product is non cytotoxic.

Latex Free: Product is latex free.

This document has been produced electronically and is valid without a signature

Date of issue: **June 2023**