
Certificate of Conformity

Product: Nunc Cryo Color Codes, yellow**Date:** 2023-11-11**Page:** 1**Item number:** 355077**Lot number:** 180889**Date of manufacture:** November 2023**Date of expiry:** October 2028

Thermo Fisher Scientific confirms that the product supplied has been inspected and tested in accordance to relevant specifications

Test Description	Requirements	Result
Visual	According to specification	Complies
Test of function	According to specification	Complies

Quality assurance of the product is performed in accordance with the requirements of the Quality Management System ISO 9001:2015 "Quality management system - Requirements" and ISO 13485:2016 "Medical devices - Quality management systems - System requirements for regulatory purposes".

Roskilde Site

Anita Kvistgaard
Operations Quality Manager

Please note that this is an electronically generated signature