

Certificate of Conformity

VWR Catalogue Number	See List
Description	Petri dishes
Country of Origin	Manufactured in Italy
Date of Issue (yyyy-mm-dd)	2024-10-16

Aseptic

Catalogue Number	Version	Diameter [mm]	Height [mm]	Packsize	Dishes Per Sleeve
390-1375	Without vents	55	14	1080	10
390-1373	With 3 vents	55	14	1080	10
391-0559	Without vents	90	14	480	20
391-0598	Without vents	90	14	720	30
391-0561	Without vents, double-outer bag	90	14	480	20
391-0565	With 1 vent	90	14	480	20
391-0556	With 3 vents	90	14	480	20
391-0582	With 3 vents	90	14	720	30
391-0603	With 3 vents, double-outer bag	90	14	720	30
391-0467	Without vents	90	16	480	20
391-0604	Without vents	90	16	720	30
391-0606	With 1 vent	90	16	720	30
391-0468	With 3 vents	90	16	480	20
391-0608	With 3 vents	90	16	720	30
391-0492	With 3 vents, 2 compartments	90	16	480	20
391-0616	Without vents	140	21	140	5
391-0618	With 3 vents	140	17.5	140	5
390-1525	With 3 vents easy open bag	55	14	1080	20

Beta-irradiated (SAL 10⁻⁶)

Catalogue Number	Version	Diameter [mm]	Height [mm]	Packsize	Dishes Per Sleeve
390-1376	Without vents	55	14	1080	10
390-1374	With 3 vents	55	14	1080	10
390-1377	Contact dish, 3 vents, double-outer bag	65	15	800	10
391-0635	With 3 vents and GRIPLID	65	15	800	10
391-0560	Without vents	90	14	480	20
391-0562	Without vents, double-outer bag	90	14	480	20
391-0599	Without vents	90	14	720	30
391-0600	Without vents, double-outer bag	90	14	720	30
391-0601	With 3 vents	90	14	720	30
391-0602	With 3 vents, double-outer bag	90	14	720	30
391-0469	With 3 vents	90	16	480	20
391-0605	Without vents	90	16	720	30
391-0607	With 1 vent	90	16	720	30
391-0609	With 3 vents	90	16	720	30
391-0617	Without vents	140	21	140	5
391-0619	With 3 vents	140	17.5	140	5

Quality System Compliance

Products are manufactured under the **ISO 9001:2015** standard. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing

Representative products samples are collected and inspected in accordance with current applicable QC requirements.

Product Specifications

The product is in conformance with the requirements of the following standards and regulations:

Material Polystyrene

Sterilization Products are sterilized using Beta rays and meets SAL 10⁻⁶ with a specified dosage range of 12.5 – 19.8 kGy. (which is not applicable for Non-Sterile products)

VWR International bv, Haasrode Researchpark 2020, Geldenaaksebaan 464, B-3001 Leuven, Belgium - <http://www.vwr.com>

ATP Assay	Not applicable.
GMO	Genetically modified materials (GMO) have not been intentionally incorporated in the manufacturing of the product.
DNase & RNase Free	Not applicable.
BSE/TSE	No use of any raw material produced from or substances derived from animal origin. The manufacturing process of the product does not use any ingredient of animal origin and no material derived from or exposed to animals affected by or under quarantine for transmitting Animal Bovine Spongiform Encephalopathy/Spongiform Encephalopathy (BSE)/(TSE).
Cytotoxicity	Not applicable.
Non-Pyrogenic Statement	Not applicable.
Latex Statement	The products are not made from natural rubber latex.
BPA Statement	Bisphenol A (BPA) has not been intentionally incorporated in the manufacturing of the products.
DEHP Statement	DEHP has not been intentionally incorporated in the manufacturing of the products.
RoHS	The product does not contain lead, cadmium, mercury and hexavalent chromium, i.e. the amount of these heavy metals does not exceed 100 ppm (Cd < 0,25 ppm, Hg < 0,2 ppm, Pb < 0,5 ppm and Cr < 1ppm).
REACH Statement	No substances as listed on >Substance of Very High Concern< candidate list published 2021-07-08 are used in manufacturing of the raw materials used to produce this product. No routinely analysis is performed.
Storage Conditions	Store in a dry place, away from sources of heat and direct sunlight.
Shelf Life	60 months.

Disclaimer: VWR states that this declaration will not discharge the user from their obligation to ensure the product is suitable for the intended use. The purpose of the product is for use in laboratory only.
This document has been produced electronically and is valid without a signature

Version History

Sl.no	Version number	Description	Month & Year
1	1.0	Created General CoQ using new layout	10-2021
2	1.1	Added GMO	07-2022
3	1.2	Changed height information for 391-0618 and 391-0619	10-2024