

Certificate of Conformity

Lot Number: 21180399

Part Number: 2065E-05

Item Description: ART 100E LR RACKED ST96TIPS 10TRAYS 5PACKS 1CASE

Manufacturing Date: 04/30/2021

Irradiation Date: 05/07/2021

Sterility Assurance Level (SAL): 10^{-6}

Sterility Assurance Level for sterilized products is ensured by irradiating the product according to ANSI/AAMI/ISO 11137- 2015 guidelines.

Molecular BioProducts Inc., a part of Thermo Fisher Scientific Inc., hereby certifies that these products are manufactured in a controlled environment in accordance with established manufacturing guidelines, quality requirements and product specifications. They are certified, by validation of the manufacturing processes, to be free of the following contaminants:

- Human and mouse DNA (Deoxyribonucleic Acid) : < 30 pg
- Ribonuclease (RNase): < 10^{-9} Kunitz Units/ μ l
- Deoxyribonuclease (DNase): < 10^{-7} Kunitz Units/ μ l
- PCR Inhibitors: not present
- Endotoxin (Pyrogen) < 0.06 EU/ml
- Adenosine-5'-triphosphate (ATP) < 10^{-13} mg/ μ l

Certified by: _____



12/04/2023

Ana Ibarria
Interim Quality Manager