

Certificate of Quality

REF

781185

LOT

B201034Y

Product description: MICROPLATE, 384 WELL, PS, F-BOTTOM,
CLEAR, LID, STERILE, SINGLE PACKED

Following inspections were carried out and passed:

raw material	☒	passed
visual	☒	passed
physical	☒	passed

Free of detectable DNase/RNase/human DNA
Representative product was tested by Real-Time-PCR.

Detection limits:

DNase 1.3×10^{-8} Kunitz units
RNase 5.0×10^{-11} Kunitz units
hDNA 5 pg

Free of detectable endotoxins (pyrogens)
Representative product was tested according to the United States Pharmacopoeia (USP) Chapter <85>
Detection limit: 0.03 EU/ml

We hereby certify that our products were manufactured and tested in accordance with the following international standards and meet the required specifications:

ISO 11137 "Sterilization of health care products - Radiation"

We are certified according to the Quality System Standard ISO 9001 and ISO 13485.



2025-10-04

STERILESAL 10^{-3}

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided.



Certificate creation date: 2020-12-06

www.gbo.com