

CERTIFICATE OF STERILITY AND QUALITY ASSURANCE

PRODUCT	LOT NUMBER	EXPIRY DATE	CERTIFICATION DATE
MW702-20	114120	31/01/2030	29/01/2025

Microloop™ Plastic Needle

PARAMETER	SPECIFICATION	RESULT
Sterilisation Method	Irradiation (Procedure conforms to ISO 11137)	PASSED
Visual inspection	Moulding must be completely smooth and without defect	PASSED

Product has been manufactured in accordance with ISO 13485:2016

Product has been manufactured in accordance with the requirements of the In Vitro Diagnostic Directive 98/79/EC 1998, Official Journal of European Communities, L331/1

For and on behalf of Medical Wire & Equipment



Quality Assurance