

Certificate of Compliance

Catalogue Number: 189171

Lot Number:

6	1	0	0	1	0	6
---	---	---	---	---	---	---

Date of Manufacture 01 May 2023

Use by Date:- 01/05/2028

Product description: 7ml Container-Ps-No Label, Plastic Cap
STERILE

Component Materials: -

Container -Polystyrene
Cap -Polypropylene

Sterilisation: -

We hereby certify that this product been sterilized with Ethylene Oxide and found to

conform to the sterility requirements of EN11135 and that the goods have been manufactured aseptically in a clean room environment as per EN ISO 13485 standards. Manufactured from Polystyrene with a Polypropylene flow seal cap and EO sterility.

Leak-proof tested in accordance with EN14245, Annex D and BS5213

This manufacturing lot has been sampled, physically examined tested, packed, stored and transported in accordance with Standard Operating Procedures as contained in the Greiner Bio-One quality manual relating to our ISO9001:2015 registration.



Noel Ennett
Operations Director