



CERTIFICATE OF MANUFACTURE/ANALYSIS

Product Size:	<u>3.78 Liters</u>	Product Name:	<u>6% Hydrogen Peroxide WFI Sterile Solution</u>
Manufacturing Date:	<u>August 4, 2020</u>	Product ID:	<u>1S05-08WR</u>
Expiration Date	<u>August 1, 2022</u>	Lot No.:	<u>308742</u>

I. The product listed above is made per the current issue of the following documentation

Raw Material Quality Control Specifications
Finished Product Master Formula
Manufacturing Instructions
Packaging Specifications
Finished Product Quality Control Specifications

Copies of this documentation are on file at STERIS Corporation, St. Louis Operations and will be provided as required to the applicable regulatory agencies.

II. The product listed above has been analyzed. The release specifications and results are shown below.

<u>Test Name</u>	<u>Method No.</u>	<u>Result</u>	<u>Specification</u>
Color	110	Colorless	Colorless
Clarity	107	Clear	Clear
pH	106	3.9	3.6 - 4.6
Hydrogen Peroxide Assay	252	6.0%	5.7 -6.3%
Specific Gravity (@25°C)	114	1.021	1.016 -1.027
Sterility Test	Current USP	Sterile	Sterile

III. The product listed above was filtered using a 0.2 micron filter prior to filling in the final package/container in an ISO Class 5 area.

Based on these results the product is released for shipment to all STERIS Corporation distribution points.

Signed:

Egisson
Quality Representative

Date: September 10, 2020