

Certificate of Conformity

VWR Inoculating Loop 1 µl, Rigid, In Bags of 10

VWR International certifies that the specified products meet all VWR specifications and quality criteria for release. The products comply in every respect with the relevant particular specifications, drawings, relevant standards and regulations in force.

VWR International certifie que les produits mentionnés respectent toutes les spécifications et critères de qualité VWR. Les produits satisfont aux spécifications particulières, aux plans, aux normes et règlements en vigueur s'y rapportant.

VWR International bestätigt hiermit, daß die angegebenen Produkte alle VWR Spezifikationen und Qualitätskriterien zur Freigabe erfüllen. Die Produkte entsprechen in jeder Hinsicht den maßgeblichen Spezifikationen, den Zeichnungen sowie den relevanten Normen und gültigen Vorschriften.

VWR Catalog No: 612-9351

Lot No: 2012047

Expiry Date: 31st May 2023

Sterility: Products labelled "STERILE R" have been gamma irradiated.

Sterilisation procedure is validated in conformity with BS EN ISO 11137-1:2015 AND BS EN ISO 11137-2:2015 to guarantee SAL 10⁻⁶.

Goods manufactured according to ISO 13485:2016 system

This document has been produced electronically and is valid without a signature

Date of issue: **August 2020**