INSTALLATION GUIDE

RephiDuo P Pack

RephiLe Catalogue Number:
RR400CP00
Corresponding Millipore Part Number:
PR0G00001

Suitable Millipore Systems:
RiOs 5 - 16
Elix 3 - 10

This product is comparable to the corresponding Millipore cartridge. Please refer to Millipore's User Manual for a detailed instruction on installation.

A Brief description of the installation process is described below:

1. Put the system into "STANDBY" mode by pressing "OPERATE/STANBY" button for two seconds. The "STANDBY" display flashes to indicate depressurization of the system.

2. Once the "STANDBY" display stops flashing:
    - Bring the pack adapter cover to the top position,
    - Remove the metal locking clip,
    - Pull out the old pack.
3. Remove the two blue protective caps on the new cartridge. Wet the two O-rings on the cartridge with a few drops of pure water.
4. Push the new cartridge so that the pack adapter metal rod goes through the hole at the top of the cartridge.
    - Lift up the cartridge slightly in order to push the bottom of it into the slot at the bottom of the system.
    - Push the top of the cartridge towards the system until it is secure.
5. Lock the cartridge in place with the metal locking clip on the end of the metal guide pin.
6. Bring the adapter cap down to its lowest position so that it covers the top of the cartridge.
7. Place the system back into "OPERATE" mode by pressing the "OPERATE/STANBY" button for two seconds. The system will automatically perform a flush and return to the normal production mode.

The system is ready for use.





RephiSolo, RephiDuo and RephiQuatro are trademarks of RephiLe Bioscience, Ltd.
© 2010. RephiLe Bioscience, Ltd. All rights reserved.
All other trademarks are trademarks of their respective manufacturers.