

Corning Incorporated - Life Sciences
6507 South 400 West
Salt Lake City UT 84107 USA
www.corning.com/lifesciences
Refer to website for regional contact information.

Page: 1 / 1

Product Name	: CLUSTER TUBE, 1.2ML, 8 STRIP, PP, NS, 120/CS	Manufacture Date	: 2019-11-18
Catalog Number	: 4408		
Lot ID	: 32219000		
Expiration Date	: 2022-12-18		


Quality Management System - Complies with the current version of the EN ISO 13485 Standard.

DNase/RNase Free - Tested by nuclease assay method and found to be free of detectable DNase/RNase contamination. The assay detection limit is 10^{-7} Kunitz units/uL for DNase and 10^{-9} Kunitz units/uL for RNase.

Quality Control Testing - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass
Packaging Inspection - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Emily D Roberts
Quality Manager