

Certificate of Conformance and Irradiation

This certification is provided as full assurance that the following product code and lot number were manufactured in accordance via a Quality System certified to ISO 9001 with prescribed procedures and specifications.

Product Code: BKTBAG

Lot Number: 014043

Manufacturing Date: 07/04/2019

Expiration Date: 07/2021

Irradiation

Radiation Lot: 19073103

Irradiation Dose Specified: 25 kGy to 50 kGy

Irradiation Dose Delivered: 26.8 kGy to 37.2 kGy

Quality Department
Contec Cleanroom Technology (Suzhou) Co., Ltd.
2019-08-08



苏州中核华东辐照有限公司
Suzhou CNNC Huadong Radiation Co., Ltd.

文件编号 (File No.): CHF/QR-8.2.4-05

辐照证明书
CERTIFICATE OF IRRADIATION

执照登记号 (License No.): 913205097186218342

Serial No.: KT190042

客户名称 (Customer): 康泰无尘科技 (苏州) 有限公司

辐照日期 (Radiation Date): 2019/07/31 灭菌批号 (Radiation Lot): 19073103

产品名称 Description	订单号 Order No.	产品货号 (代号) Stock No. (Code No.)	产品批号 Lot No.	箱数 Cases
无尘擦拭布	N/A	GCQT0150	014111	96
无尘包装袋	N/A	BKT BAG	014043	512
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A	N/A

客户要求剂量 (Sterilization Dose Required by Customer):

最低剂量 (Minimum Dose): 25.0 kGy 最高剂量 (Maximum Dose): 50.0 kGy

苏州中核华东辐照有限公司特此证明上述产品已经钴-60射线辐照。通过已评估测量不确定度 (5%) 的剂量测量系统测量, 吸收剂量如下:

Suzhou CNNC Huadong Radiation Co., Ltd. hereby certifies that above products have been irradiated by cobalt-60 gamma ray. The absorbed dose measured through system of evaluated uncertainty (5%) is as follows:

剂量计名称 (Dosimeter Name): 重铊酸钾 (银) 剂量计批号 (Dosimeter Lot): 181217

产品实际最低吸收剂量 (Minimum Absorbed Dose): 26.8 kGy

产品实际最高吸收剂量 (Maximum Absorbed Dose): 37.2 kGy

签发 (Issued by):

日期 (Date): 2019.8.1

本公司通过了ISO13485体系和ISO11137标准的认证 Our company is certified for conformance to ISO13485 and ISO11137 standards by TUV ps
地址: 江苏省苏州市吴江区交通路4756号 Address: #4756, Jiaotong rd., Wujiang District, Suzhou City, Jiangsu Province, PRC