

Certificate of Conformity

Product Code: LW4010

Item Description: Pipette Graduated 1ml Sterile Pastette® Pack of 25x20

Lot Number: 25090521

Material: Low Density Polyethylene

Sterility: Sterile

Sterilisation Method: E-beam Irradiation

Sterility Details: Sterility Assurance Level: 10^{-6}

Where a Sterility Assurance Level is stated, it is ensured by irradiating the product according to ANSI/AAMI/ISO 11137-2006 guidelines.

Date of Expiry: 31-MAR-29

Manufacturing Location: MEXICO

Other information relating to this product:

Length: 155mm

Bulb Draw: 3.4ml

Drops per ml: 25

Drop Volume: 40µl

Graduations at 0.25ml

Alpha Laboratories Ltd hereby certifies that these products have been manufactured in a controlled environment in accordance with established manufacturing guidelines. All items have been inspected and meet packaging requisites.

Certified by:



Michelle Holland, Managing Director

Date: 13 March 2026



Alpha Laboratories Ltd
40 Parham Drive
Hampshire SO50 4NU

Tel: +44 (0)23 8048 3000
Email: alpha@alphalabs.co.uk
Web: www.alphalabs.co.uk