



Certificate of Conformance and Sterility

This certification is provided as full assurance that the following product code and lot number were manufactured in accordance via a Quality System certified to ISO 9001 with prescribed procedures and specifications.

Product Code:	PS-7030IR
Lot Number:	778195-0-0-1
Manufacturing Date:	08/13/2020
Isopropyl Alcohol:	70 ± 3% by volume with the remainder DI H2O
Expiration Date:	08/2022

Sterility

Gamma Processing Run ID:	2722758
Irradiation Dose Specified:	18.20 kGy to 40 kGy
Irradiation Dose Delivered:	21.4 kGy to 31.8 kGy
Validated Sterile Product:	Validation Sterile packets available upon request.
Sterility Assurance Level:	10 ⁻⁶ (SAL determined by test method AAMI/ISO 11137)

The expiration date is the date of manufacture plus two years. If you have any questions please contact us.

QC Supervisor
coc@contecinc.com

Friday, August 21, 2020

NORTH AMERICA

Contec, Inc.
525 Locust Grove
Spartanburg, SC 29303 USA
tel: +1 864 503 8333
fax: +1 864 503 8444

EUROPE

Contec, Inc.
ZI du Prat - R.P. 3707
F-56037 VANNES Cedex France
tel: +33 (0)2 97 43 76 98
fax: +33 (0) 97 43 76 86

ASIA

Contec Cleanroom Technology (Suzhou) Co., Ltd.
17 Longyun Road
Suzhou Industrial Park, Suzhou 215024 China
tel: +86 512 6274 4050
fax: +86 512 6274 4051

web: www.contecinc.com
e-mail: info@contecinc.com

Page 1 of 1

DID: 2186B



Certificate of Processing

STERIGENICS 10811 Withers Cove Park Dr. Charlotte NC 28278
TEL 704 588-6877 FAX 704 588-8355 www.sterigenics.com

R55480102

08/20/2020 16:03:54 GMT
Page 1 of 1

Customer Name:	Contec- PO Box 530	Processing Facility:	Charlotte Gamma	Work Order #	2722758
P.O.#	111626			Sales Order #	2570070
18.2-40.0 kGy		CNT4, Gamma Treatment	Irradiation Date/Time:	08/19/2020 17:03:00 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	560	CA	PS-7030IR	Clean Room Product	778195-0-0-1	111626
	560	CA	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	18.2 kGy	40.0 kGy	21.4 KGY	Pass	BTOWERY	08/20/2020 01:13:19 GMT
		Reason Code Test				Benjamin Towery	
450.00	Maximum Dose	18.2 kGy	40.0 kGy	31.8 KGY	Pass	BTOWERY	08/20/2020 01:13:35 GMT
		Reason Code Test				Benjamin Towery	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: BRITTANY LONG
Reason: Work Order Completions

Date: 08/20/2020 16:00:06 GMT