

Certificate of Quality

REF

786201

LOT

B20073XK

Product description: MASTERBLOCK®, 96 WELL, 0.5 ML, V-BOTTOM,
NATURAL, 8 PCS./BAG

Following inspections were carried out and passed:

raw material	☒	passed
visual	☒	passed
physical	☒	passed

Free of detectable DNase/RNase/human DNA
Representative product was tested by Real-Time-PCR.

Detection limits:

DNase 1.3×10^{-8} Kunitz units
RNase 5.0×10^{-11} Kunitz units
hDNA 5 pg

Free of detectable endotoxins (pyrogens)
Representative product was tested according to the United States Pharmacopoeia (USP) Chapter <85>
Detection limit: 0.03 EU/ml

We are certified according to the Quality System Standard ISO 9001 and ISO 13485.

This certificate is valid for the above mentioned lot and has no validity if no lot number is provided.



Certificate creation date: 2020-08-31

www.gbo.com