

Quality Certificate

VWR® POWDER-FREE NITRILE EXAMINATION GLOVES

As per the manufacturer, the below product meets the following criteria:

North American Catalog Number:	82026-428
Lot Number:	PT18J206
Description:	VWR Glove Nitrile Pf L Pk100
Manufacturing Date:	July 2018
Product Shelf Life:	June 2023
Country of Origin:	Indonesia

Quality System Compliance:

Products are manufactured under FDA510K, ISO 9001:2008, ISO 13485:2003I, CMDCAS Quality standard and Good Manufacturing Practices. Products are inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP.

QC Testing: Representative production samples are collected and inspected in accordance with current applicable Product specification: Meet or exceed ASTM D6319 and ISO 2859

Accuracy:	Water tight test:	Inspection level = G1;	AQL 1.5
	Visual inspection:	Inspection level = G1;	AQL 1.5 (critical defects) AQL 2.5 (major defects) AQL 4.0 (minor defects)
	Dimensional:	Inspection level = S2;	AQL 4.0
	Physical Property:	Inspection level = S2;	AQL 4.0
	Powder Content:	Inspection level = 5 pieces;	Average ≤ 2 mg/glove
	Moisture Content:	Inspection level = 5 pieces;	Average ≤ 1%
	Glove count:	Inspection level = 5 dispensers;	Tolerance 100+/-2 piecesz

Product Specification: This product is Class I Category device as defined by the FDA in 21CFR parts 807 sub parts C

Material: Nitrile Butadiene Rubber
BSE-TSE: N/A

Non-Pyrogenic: N/A

RNase/DNase Free: N/A

Cytotoxicity: N/A

Latex Statement: Latex free

Signed:



Jamie Ethier
VP Global Quality
VWR, Part of Avantor

Date: January 11, 2020