Bel-Art Products
Wayne, New Jersey 07470
Tel: 973-694-0500 * Fax: 973-694-7199
Internet Address: http://www.bel-art.com

Technical Data Sheet

Date: January 4, 2016

Product #: 36940-0000

Description: "Sterileware" Sterile Sampler Spoons

Material: High-Impact Polystyrene Natural, sterilized in individual polyethylene bags. Materials meet FDA requirements for regulation 21 part 177.1640, White polystyrene Colorant meets FDA regulations 21 CFR 177.1640, 178.3297, 175.300

Dimensions: Overall 7" (178mm), Handle 6.12"
OD Bowl 0.81"(20mm)W x 0.96" (24.4mm)L x 0.26" (6.60mm) D,
ID Bowl: 0.73" (18.5mm)W x .94" (23.9mm)L x 0.22"(5.58mm) D,
Measurements are at max. point and do not include angles etc.
Volume approx-1.23ml

Other: No testing is done for Pyrogens / End toxins at this time
Based on the process and materials involved in the manufacturing of this product, this product is considered to be BSE/TSE protein free,
Product should remain in its irradiated state as long as outer packaging remains completely intact. ANSI/AAMMI/ISO 11137-1994 Method 1, guidelines are followed. Gamma Irradiated between 23.5 kGy to 50.0 kGy
Manufactured: Bel-Art Products , 2024 Broad Street, Pocomoke City MD, USA