



CERTIFICATE OF CONFORMANCE AND STERILIZATION

Date: 2-Dec-19

This certification is provided as full assurance that the following product code was manufactured in accordance with prescribed procedures and specifications.

Product/Material Number:	C2-1212IR
Product Description:	STERILESORB* C/P 12INx12IN
Lot Number:	172386
Irradiation Run Number:	2559678
Irradiation Dose Delivered:	32.5kGy to 42.1kGy
Manufacture Date:	11/2019
Expiration Date:	11/2021

The above product has been Gamma Irradiated to Sterility Assurance Level of 10^{-6} in accordance with the guidelines of the Association for the Advancement of Medical Instrumentation.

A handwritten signature in black ink, appearing to read "C. Inthavongsa".

Champagne Inthavongsa
Quality Supervisor
cinthavongsa@contecinc.com

Certificate of Processing

STERIGENICS 10811 Withers Cove Park Dr. Charlotte NC 28278
 TEL 704 588-6877 FAX 704 588-8355 www.sterigenics.com

R55480102

11/25/2019 15:04:17 GMT
 Page 1 of 1

Customer Name:	FG Clean Wipes- Chicopee, MA.	Processing Facility:	Charlotte Gamma	Work Order #	2559678
P.O.#	38160			Sales Order #	2373911
25-50 kGy		LYM2, Gamma Treatment	Irradiation Date/Time:	11/23/2019 10:04:00 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	320	CA	C2-1212IR	Wiping Cloth	172386	NA
102.000	720	CA	C2-99IR/25	Wiping Cloth	172382	NA
	1040	CA	Total			

Processing Summary

Op#	Operation Name	Location	Date/Time In (GMT)	Date/Time Out (GMT)
400.00	Receive Product	UNPROC	11/18/2019 15:56:00	11/21/2019 20:17:16
410.00	Prepare Product	UNPROC	11/21/2019 20:17:16	11/22/2019 22:23:01
420.00	Load Product - A123	UNPROC	11/22/2019 22:23:01	11/23/2019 10:04:00
430.00	Process Product	UNPROC	11/23/2019 10:04:00	11/23/2019 10:10:14
440.00	Unload Product	UNPROC	11/23/2019 10:10:14	11/23/2019 21:22:40
450.00	Collect Run Data	UNPROC	11/23/2019 21:22:40	11/25/2019 13:32:59

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail	-----Signed By----- User	Date /Time
450.00	Minimum Dose	25.0 kGy	50.0 kGy	32.5 KGY	Pass	SJDAVIS	11/25/2019 13:32:01 GMT
	Reason Code Test					STEVEN DAVIS	
450.00	Maximum Dose	25.0 kGy	50.0 kGy	42.1 KGY	Pass	SJDAVIS	11/25/2019 13:32:09 GMT
	Reason Code Test					STEVEN DAVIS	

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: WENDY KOBLISH
 Reason: Work Order Completions

Date: 11/25/2019 15:03:03 GMT