



AquaPhoenix[®]
S C I E N T I F I C



CHEMetrics[®]
By AquaPhoenix Scientific

Simplicity in Water Analysis

Certificate of Conformance

Catalog No: R-7540

Product Description: Oxygen Refill, Dissolved

Lot: 4GC0937

Date of Manufacture: 28 March 2024

Date Certified: 28 March 2024

Expiration Date: June 2026

Test	Specification	Answer
Reagent Appearance	colorless	Pass
Dissolved Oxygen Test at 0 ppb	0 to 5 ppb	Pass
Dissolved Oxygen Test at 80 ppb	60 to 100 ppb	Pass

CHEMetrics by AquaPhoenix Scientific, LLC hereby certifies that the following water analysis products have been manufactured and evaluated in accordance with strict Quality Control (QC) and Quality Assurance (QA) protocols and/or conform to manufacturer specifications. Where applicable, laboratory test results as well as retention samples for each production batch are on file at CHEMetrics.

Jennifer R. Nash

Signature

Jennifer R. Nash
Director of Quality